

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137746

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/04/2025	08:30
2	Q3500-01		1	11	100.00	>212.0	>212.0	11/04/2025	11:00
3	Q3500-01DUP		1	11	100.00	>212.0	>212.0	11/04/2025	11:30
4	Q3502-01		1	12	100.00	>212.0	>212.0	11/04/2025	12:00
5	Q3502-04		1	4	54.00	129.2	129.2	11/04/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137746

WorkList Name : fp-11-3

WorkList ID : 192843

Department : Wet-Chemistry

Date : 11-03-2025 08:34:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3500-01	459	Solid	Flash Point	Cool 4 deg C	PSEG03	D41	10/30/2025	1010B
Q3502-01	CATCH-CAN-USED-OIL	Solid	Flash Point	Cool 4 deg C	PSEG03	D41	10/30/2025	1010B
Q3502-04	PARTS-CLEANER	Solid	Flash Point	Cool 4 deg C	PSEG03	D41	10/30/2025	1010B

DateTime 11/04/2025 08:15
 Raw Sample Received by: RMCWG
 Raw Sample Relinquished by: J06007

DateTime 11/04/2025 13:10
 Raw Sample Received by: J06007
 Raw Sample Relinquished by: RMCWG