

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137757

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/04/2025	13:15
2	CAL2	1	Water	NA	NA	20.2	7.00	11/04/2025	13:16
3	CAL3	1	Water	NA	NA	20.2	10.02	11/04/2025	13:17
4	ICV	1	Water	NA	NA	20.2	7.01	11/04/2025	13:20
5	CCV1	1	Water	NA	NA	20.2	2.01	11/04/2025	13:22
6	Q3368-13	1	Solid	20.02	20	20.8	7.62	11/04/2025	13:30
7	Q3368-14	1	Solid	20.03	20	20.1	8.01	11/04/2025	13:37
8	Q3368-14DUP	1	Solid	20.03	20	20.3	8.02	11/04/2025	13:39
9	CCV2	1	Water	NA	NA	20.2	12.02	11/04/2025	13:40

WORKLIST(Hardcopy Internal Chain)

UB 137757

WorkList Name : ph q3368

WorkList ID : 192867

Department : Wet-Chemistry

Date : 11-04-2025 12:57:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3368-13	OUTSIDE-DUMPSTER	Solid	pH	Cool 4 deg C	ARAM01	D41	10/16/2025	9045D
Q3368-14	PRESS-SLUDGE	Solid	pH	Cool 4 deg C	ARAM01	D41	10/16/2025	9045D

Date/Time 11/04/25 13:10

Raw Sample Received by: To (see)

Raw Sample Relinquished by: To (see)

Date/Time 11/04/25 13:50

Raw Sample Received by: To (see)

Raw Sample Relinquished by: To (see)