

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB137776

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/05/2025	10:10
2	CAL2	1	Water	NA	NA	20.3	7.00	11/05/2025	10:11
3	CAL3	1	Water	NA	NA	20.4	10.03	11/05/2025	10:14
4	ICV	1	Water	NA	NA	20.1	6.98	11/05/2025	10:15
5	CCV1	1	Water	NA	NA	20.3	2.02	11/05/2025	10:19
6	Q3534-01	1	Water	NA	NA	20.6	6.20	11/05/2025	10:25
7	Q3534-02	1	Water	NA	NA	20.4	6.52	11/05/2025	10:26
8	Q3534-03	1	Water	NA	NA	20.5	6.64	11/05/2025	10:33
9	Q3534-04	1	Water	NA	NA	20.9	6.56	11/05/2025	10:37
10	Q3534-05	1	Water	NA	NA	20.1	6.32	11/05/2025	10:40
11	Q3534-06	1	Water	NA	NA	20.8	6.56	11/05/2025	10:47
12	Q3534-07	1	Water	NA	NA	20.8	6.70	11/05/2025	10:50
13	Q3537-03	1	Water	NA	NA	20.6	6.80	11/05/2025	11:00
14	Q3541-01	1	Water	NA	NA	20.7	8.49	11/05/2025	11:10
15	Q3542-01	1	Water	NA	NA	20.6	9.70	11/05/2025	11:20
16	CCV2	1	Water	NA	NA	20.3	12.02	11/05/2025	11:25
17	Q3542-01DUP	1	Water	NA	NA	20.8	9.72	11/05/2025	11:27
18	CCV3	1	Water	NA	NA	20.2	2.01	11/05/2025	11:30

WORKLIST(Hardcopy Internal Chain)

137776

WorkList Name : ph q w3542

WorkList ID : 192887

Department : Wet-Chemistry

Date : 11-05-2025 08:39:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3534-01	MW-6	Water	pH	Cool 4 deg C	REMI02	D31	11/03/2025	9040C
Q3534-02	MW-1	Water	pH	Cool 4 deg C	REMI02	D31	11/03/2025	9040C
Q3534-03	MW-11	Water	pH	Cool 4 deg C	REMI02	D31	11/03/2025	9040C
Q3534-04	MW-5	Water	pH	Cool 4 deg C	REMI02	D31	11/04/2025	9040C
Q3534-05	MW-EB-1	Water	pH	Cool 4 deg C	REMI02	D31	11/03/2025	9040C
Q3534-06	MW-EB-2	Water	pH	Cool 4 deg C	REMI02	D31	11/04/2025	9040C
Q3534-07	MW-EB-3	Water	pH	Cool 4 deg C	REMI02	D31	11/03/2025	9040C
Q3537-03	MW-17-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	11/04/2025	9040C
Q3541-01	N48969	Water	pH	Cool 4 deg C	PSEG03	D41	11/04/2025	9040C
Q3542-01	AUD-25-182-194	Water	pH	Cool 4 deg C	PSEG03	D41	11/04/2025	9040C

Date/Time 11/05/25 08:45

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 11/05/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]