

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137777

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/05/2025	10:00
2	CAL2	1	Water	NA	NA	20.2	7.00	11/05/2025	10:01
3	CAL3	1	Water	NA	NA	20.3	10.02	11/05/2025	10:02
4	ICV	1	Water	NA	NA	20.1	7.00	11/05/2025	10:05
5	CCV1	1	Water	NA	NA	20.2	2.01	11/05/2025	10:06
6	Q3539-01	1	Solid	20.02	20	23.5	8.26	11/05/2025	10:15
7	Q3543-01	1	Solid	20.03	20	23.6	7.05	11/05/2025	10:25
8	Q3543-01DUP	1	Solid	20.04	20	23.7	7.07	11/05/2025	10:27
9	CCV2	1	Water	NA	NA	20.2	12.02	11/05/2025	10:30

WORKLIST(Hardcopy Internal Chain)

VB 137777

WorkList Name : ph s q3473 WorkList ID : 192888 Department : Wet-Chemistry Date : 11-05-2025 08:41:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3539-01	JF-HUEBNER	Solid	pH	Cool 4 deg C	AMBR01	D41	11/04/2025	9045D
Q3543-01	☒ TRE-25-0073	Solid	pH	Cool 4 deg C	PSEG03	D41	11/04/2025	9045D

Date/Time 11/05/25 08:45
 Raw Sample Received by: SS Woy
 Raw Sample Relinquished by: SS Woy

Date/Time 11/05/25 14:30
 Raw Sample Received by: SS Woy
 Raw Sample Relinquished by: SS Woy