

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB137787

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	11/05/2025	11:00
2	Q3517-02		1	13	100.00	>212.0	>212.0	11/05/2025	11:30
3	Q3517-02DUP		1	13	100.00	>212.0	>212.0	11/05/2025	12:00
4	Q3517-03		1	12	100.00	>212.0	>212.0	11/05/2025	12:30
5	Q3518-01		1	12	100.00	>212.0	>212.0	11/05/2025	13:00
6	Q3518-02		1	14	100.00	>212.0	>212.0	11/05/2025	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB137787

WorkList Name : FP-11-04

WorkList ID : 192866

Department : Wet-Chemistry

Date : 11-04-2025 10:08:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3517-02	NWB-2156	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/31/2025	1010B
Q3517-03	NWB-2226	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/31/2025	1010B
Q3518-01	50771	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/31/2025	1010B
Q3518-02	50831	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/31/2025	1010B

Date/Time 11/05/25 10:15
 Raw Sample Received by: 12600
 Raw Sample Relinquished by: 12600

Date/Time 11/05/25 16:15
 Raw Sample Received by: 12600
 Raw Sample Relinquished by: 12600