

LB137791


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34430



A= 0.0000e+000

B= 2.9017e-004

C= 3.2888e-002

Rho= 0.9999344

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	-87	0.000	-87	0			
0.2	0.200	0.154	-0.046	416	0.0 %	416	0			
2.5	2.500	2.565	0.065	8725	0.0 %	8725	0			
5.0	5.000	5.011	0.011	17156	0.0 %	17156				
7.5	7.500	7.440	-0.060	25525	0.0 %	25525				
10.0	10.000	10.024	0.024	34430	0.0 %	34430				

LB137791 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Nov 2025 07:27:52

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-87	PPB	Std	0.0000	7470A	06 Nov 2025 07:44:58
S	0.2 - 1	50.2	-	416	PPB	Std	0.2000	7470A	06 Nov 2025 07:47:16
S	2.5 - 1	52.5	-	8725	PPB	Std	2.5000	7470A	06 Nov 2025 07:49:33
S	5.0 - 1	55	-	17156	PPB	Std	5.0000	7470A	06 Nov 2025 07:51:50
S	7.5 - 1	57.5	-	25525	PPB	Std	7.5000	7470A	06 Nov 2025 07:54:09
S	10.0 - 1	511	-	34430	PPB	Std	10.0000	7470A	06 Nov 2025 08:01:58
U	ICV100 - 1	ICV100	4.2061	14382	PPB	SMPL	-7470A		06 Nov 2025 08:06:44
U	ICB100 - 1	ICB100	-0.0182	-176	PPB	SMPL	-7470A		06 Nov 2025 08:09:06
U	CCV31 - 1	CCV31	4.9873	17074	PPB	SMPL	-7470A		06 Nov 2025 08:11:26
U	CCB31 - 1	CCB31	-0.0307	-219	PPB	SMPL	-7470A		06 Nov 2025 08:13:41
U	CRA - 1	CRA	0.2337	692	PPB	SMPL	-7470A		06 Nov 2025 08:18:37
U	HighStd - 1	HighStd	9.8793	33933	PPB	SMPL	-7470A		06 Nov 2025 08:20:52
U	ChkStd - 1	ChkStd	6.9680	23900	PPB	SMPL	-7470A		06 Nov 2025 08:23:07
U	PB170407BL - 1	PBW	-0.0608	-323	PPB	SMPL	-7470A		06 Nov 2025 08:25:30
U	PB170407BS - 1	LCSW	4.1678	14250	PPB	SMPL	-7470A		06 Nov 2025 08:27:51
U	Q3534-01 - 1	MW-6	0.0146	-63	PPB	SMPL	-7470A		06 Nov 2025 08:30:07
U	Q3534-02 - 1	MW-1	0.1194	298	PPB	SMPL	-7470A		06 Nov 2025 08:32:34
U	Q3534-03 - 1	MW-11	0.0256	-25	PPB	SMPL	-7470A		06 Nov 2025 08:34:59
U	Q3534-04 - 1	MW-5	0.0459	45	PPB	SMPL	-7470A		06 Nov 2025 08:37:33
U	Q3534-05 - 1	MW-EB-1	0.1231	311	PPB	SMPL	-7470A		06 Nov 2025 08:39:48
U	CCV32 - 1	CCV32	5.2043	17822	PPB	SMPL	-7470A		06 Nov 2025 08:42:03
U	CCB32 - 1	CCB32	-0.0347	-233	PPB	SMPL	-7470A		06 Nov 2025 08:44:19
U	Q3534-06 - 1	MW-EB-2	0.0671	118	PPB	SMPL	-7470A		06 Nov 2025 08:46:37
U	Q3534-07 - 1	MW-EB-3	0.0956	216	PPB	SMPL	-7470A		06 Nov 2025 08:48:53
U	Q3534-09 - 1	FB-1	0.0036	-101	PPB	SMPL	-7470A		06 Nov 2025 08:51:08
U	Q3537-03 - 1	MW-17-PURGE-WATER	0.0291	-13	PPB	SMPL	-7470A		06 Nov 2025 08:53:24
U	Q3541-01 - 1	N48969	0.0674	119	PPB	SMPL	-7470A		06 Nov 2025 08:55:40
U	Q3542-01 - 1	AUD-25-182-194	0.0642	108	PPB	SMPL	-7470A		06 Nov 2025 08:57:56
U	Q3542-01DUP - 1	AUD-25-182-194DUP	0.0523	67	PPB	SMPL	-7470A		06 Nov 2025 09:00:12
U	Q3542-01MS - 1	AUD-25-182-194MS	3.9969	13661	PPB	SMPL	-7470A		06 Nov 2025 09:02:29
U	Q3542-01MSD - 1	AUD-25-182-194MSD	4.1020	14023	PPB	SMPL	-7470A		06 Nov 2025 09:04:47
U	CCV33 - 1	CCV33	5.1074	17488	PPB	SMPL	-7470A		06 Nov 2025 09:07:40
U	CCB33 - 1	CCB33	-0.0553	-304	PPB	SMPL	-7470A		06 Nov 2025 09:10:08
U	Q3542-01LX5 - 1		0.0442	39	PPB	SMPL	-7470A		06 Nov 2025 09:12:27
U	Q3542-01A - 1		4.2183	14424	PPB	SMPL	-7470A		06 Nov 2025 09:14:43
U	CCV34 - 1	CCV34	5.2954	18136	PPB	SMPL	-7470A		06 Nov 2025 09:16:58
U	CCB34 - 1	CCB34	0.0184	-50	PPB	SMPL	-7470A		06 Nov 2025 09:19:27