

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137796

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	28.00	82.4	81.7	11/06/2025	10:05
2	Q3537-03		1	14	100.00	>212.0	>212.0	11/06/2025	10:35
3	Q3537-03DUP		1	14	100.00	>212.0	>212.0	11/06/2025	11:05
4	Q3541-01		1	13	100.00	>212.0	>212.0	11/06/2025	11:35
5	Q3542-01		1	15	100.00	>212.0	>212.0	11/06/2025	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16137796

WorkList Name: FP-11-05

WorkList ID: 192944

Department: Wet-Chemistry

Date: 11-05-2025 12:05:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3537-03	MW-17-PURGE-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	D41	11/04/2025	1010B
Q3541-01	N48969	Water	Flash Point	Cool 4 deg C	PSEG03	D41	11/04/2025	1010B
Q3542-01	AUD-25-182-194	Water	Flash Point	Cool 4 deg C	PSEG03	D41	11/04/2025	1010B

Date/Time 11/06/2025 09:55
 Raw Sample Received by: EM CWC
 Raw Sample Relinquished by: EM CWC

Date/Time 11/06/2025 12:45
 Raw Sample Received by: EM CWC
 Raw Sample Relinquished by: EM CWC