

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB137818

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP115524
Turbidity Calibration std, 0NTU	WP115522
Turbidity Calibration std, 20NTU	WP115517
Turbidity Calibration std, 0.5NTU	WP115521
Turbidity Calibration std, 40NTU	WP115516
Turbidity Calibration std, 1NTU	WP115520
Turbidity Calibration std, 5NTU	WP115519
Turbidity LOQ std, 1.0NTU	WP115523
Turbidity Calibration - CCV std, 10 NTU	WP115518
10 NTU Standard 500 ml	W3116

Intercept: -0.019

Slope: 1.0544

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.16		11/07/2025	08:32
2	CAL2	1	1	1.140	1.10	9.9	11/07/2025	08:35
3	CAL3	5	1	5.260	5.01	0.1	11/07/2025	08:38
4	CAL4	10	1	9.870	9.38	-6.2	11/07/2025	08:41
5	CAL5	20	1	21.500	20.41	2	11/07/2025	08:44
6	CAL6	40	1	42.100	39.95	-0.1	11/07/2025	08:47



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB137818

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.600	10.071	11/07/2025	08:50
2	ICB		1	0.152	0.162	11/07/2025	08:53
3	CCV1	10	1	9.730	9.246	11/07/2025	08:56
4	CCB1		1	0.161	0.171	11/07/2025	08:59
5	LB137818BL		1	0.168	0.177	11/07/2025	09:02
6	Q3530-07		1	0.422	0.418	11/07/2025	09:05
7	Q3530-08		1	0.968	0.936	11/07/2025	09:08
8	CCV2	10	1	9.980	9.483	11/07/2025	09:30
9	CCB2		1	0.152	0.162	11/07/2025	09:33