

66137870

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

11/12/2025 13:57

Test: TKN-NH3

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	5.090	0.0	1.016	
ICB1	-0.014	0.0	0.021	
CCV1	4.981	0.0	0.995	
CCB1	0.011	0.0	0.025	
RL CHECK	0.475	0.0	0.116	
11/12/25 RM PB170452BL 510b1	-0.016	0.0	0.020	95% (50-150)
11/12/25 RM PB170452BS 510b3	5.117	0.0	1.021	11/12/2025 RM
Q3530-07	0.241	0.0	0.070	
Q3530-08	0.537	0.0	0.128	
Q3552-01	32.782	0.0	6.418	Abs. high, Init abs., Tes
Q3552-02	13.160	0.0	2.590	Init abs., Test limit hig
Q3552-03	33.601	0.0	6.578	Abs. high, Init abs., Tes
Q3552-04	2.872	0.0	0.583	
Q3552-04DUP	2.701	0.0	0.550	
CCV2	5.046	0.0	1.008	
CCB2	0.028	0.0	0.029	
Q3552-04MS	7.307	0.0	1.449	
Q3552-04MSD	7.196	0.0	1.427	
Q3552-05	31.018	0.0	6.074	Init abs., Test limit hig
Q3560-01	8.250	0.0	1.633	
Q3560-03	0.101	0.0	0.043	
Q3560-05	0.404	0.0	0.102	
Q3560-07	29.862	0.0	5.849	Init abs., Test limit hig
Q3560-09	1.441	0.0	0.304	
Q3569-01	27.813	0.0	5.449	Init abs., Test limit hig
Q3569-02	30.898	0.0	6.051	Init abs., Test limit hig
CCV3	5.068	0.0	1.012	
CCB3	-0.014	0.0	0.021	
Q3584-01	0.655	0.0	0.151	
Q3584-02	0.747	0.0	0.169	
Q3584-03	2.170	0.0	0.447	
Q3584-07	31.572	0.0	6.182	Init abs., Test limit hig
PB170482BL	0.378	0.0	0.097	
PB170482BS	5.351	0.0	1.067	
Q3483-07	30.915	0.0	6.054	Init abs., Test limit hig
Q3530-01	0.232	0.0	0.069	
Q3530-02	0.514	0.0	0.124	
Q3586-02	19.729	0.0	3.872	Init abs., Test limit hig
CCV4	5.118	0.0	1.022	
CCB4	0.161	0.0	0.055	
Q3586-02DUP	19.660	0.0	3.858	Init abs., Test limit hig
Q3586-02MS	22.223	0.0	4.358	Init abs., Test limit hig
Q3586-02MSD	22.373	0.0	4.388	Init abs., Test limit hig
Q3552-01DLX20	7.902	0.0	1.565	
Q3552-02DLX2	6.532	0.0	1.298	
Q3552-03DLX10	8.534	0.0	1.688	
Q3552-05DLX10	6.661	0.0	1.323	
Q3560-07DLX10	4.854	0.0	0.970	
Q3569-01DLX10	3.594	0.0	0.724	
CCV5	4.965	0.0	0.992	
CCB5	-0.023	0.0	0.019	
Q3569-02DLX10	7.098	0.0	1.408	
Q3483-07DLX10	5.586	0.0	1.113	
Q3586-02DLX5	4.526	0.0	0.906	
Q3586-02DUPDLX5	4.428	0.0	0.887	

6137870

Test results

Aquakem 7.2AQ1

Page: 2

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

11/12/2025 13:57

Reviewed by : RM

Instrument ID : Konelab

Test: TKN-NH3

Sample Id	Result	Dil. 1 +	Response	Errors
Q3484-07DLX20	6.876	0.0	1.365	
CCV6	5.182	0.0	1.034	
CCB6	0.015	0.0	0.026	

N	58
Mean	8.526
SD	10.5347
CV%	123.57

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 12 10:23:27 2025

Wed Nov 12 13:45:39 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	TKN-NH3	P	-0.013	mg/l	11/12/2025 10:23:27	
0.5PPM	A	TKN-NH3	P	0.4527	mg/l	11/12/2025 10:23:28	
1.0PPM	A	TKN-NH3	P	0.9186	mg/l	11/12/2025 10:23:29	
2.5PPM	A	TKN-NH3	P	2.5078	mg/l	11/12/2025 10:23:30	
5.0PPM	A	TKN-NH3	P	5.1098	mg/l	11/12/2025 10:23:31	
6.7PPM	A	TKN-NH3	P	6.8781	mg/l	11/12/2025 10:23:32	
10.0PPM	A	TKN-NH3	P	9.8128	mg/l	11/12/2025 10:23:33	
ICV1	S	TKN-NH3	P	5.0902	mg/l	11/12/2025 11:32:42	
ICB1	S	TKN-NH3	P	-0.0135	mg/l	11/12/2025 11:32:44	
CCV1	S	TKN-NH3	P	4.9809	mg/l	11/12/2025 11:32:46	
CCB1	S	TKN-NH3	P	0.0107	mg/l	11/12/2025 11:32:49	
RL CHECK	S	TKN-NH3	P	0.475	mg/l	11/12/2025 11:32:51	
PB170510BL	S	TKN-NH3	P	-0.0164	mg/l	11/12/2025 11:32:53	
PB170510BS	S	TKN-NH3	P	5.1167	mg/l	11/12/2025 11:43:23	
Q3530-07	S	TKN-NH3	P	0.2412	mg/l	11/12/2025 11:43:26	
Q3530-08	S	TKN-NH3	P	0.5368	mg/l	11/12/2025 11:43:27	
Q3552-01	S	TKN-NH3	P	32.7823	mg/l	11/12/2025 11:43:29	
Q3552-02	S	TKN-NH3	P	13.1601	mg/l	11/12/2025 11:43:30	
Q3552-03	S	TKN-NH3	P	33.6011	mg/l	11/12/2025 11:43:31	
Q3552-04	S	TKN-NH3	P	2.8716	mg/l	11/12/2025 11:43:33	
Q3552-04DUP	S	TKN-NH3	P	2.7009	mg/l	11/12/2025 11:43:34	
CCV2	S	TKN-NH3	P	5.0457	mg/l	11/12/2025 11:54:08	
CCB2	S	TKN-NH3	P	0.0282	mg/l	11/12/2025 11:54:09	
Q3552-04MS	S	TKN-NH3	P	7.3071	mg/l	11/12/2025 11:54:10	
Q3552-04MSD	S	TKN-NH3	P	7.196	mg/l	11/12/2025 11:54:11	
Q3552-05	S	TKN-NH3	P	31.0181	mg/l	11/12/2025 11:54:14	
Q3560-01	S	TKN-NH3	P	8.2499	mg/l	11/12/2025 11:54:15	
Q3560-03	S	TKN-NH3	P	0.101	mg/l	11/12/2025 11:54:16	
Q3560-05	S	TKN-NH3	P	0.4044	mg/l	11/12/2025 11:54:17	
Q3560-07	S	TKN-NH3	P	29.8624	mg/l	11/12/2025 11:54:18	
Q3560-09	S	TKN-NH3	P	1.4411	mg/l	11/12/2025 12:04:51	
Q3569-01	S	TKN-NH3	P	27.813	mg/l	11/12/2025 12:04:52	
Q3569-02	S	TKN-NH3	P	30.8975	mg/l	11/12/2025 12:04:53	
CCV3	S	TKN-NH3	P	5.0676	mg/l	11/12/2025 12:04:54	
CCB3	S	TKN-NH3	P	-0.0141	mg/l	11/12/2025 12:04:55	
Q3584-01	S	TKN-NH3	P	0.6552	mg/l	11/12/2025 12:04:56	
Q3584-02	S	TKN-NH3	P	0.7465	mg/l	11/12/2025 12:04:57	
Q3584-03	S	TKN-NH3	P	2.1705	mg/l	11/12/2025 12:04:58	
Q3584-07	S	TKN-NH3	P	31.5722	mg/l	11/12/2025 12:04:59	

PB170482BL	S	TKN-NH3	P	0.3784 mg/l	11/12/2025 12:05:00
PB170482BS	S	TKN-NH3	P	5.3513 mg/l	11/12/2025 12:05:01
Q3483-07	S	TKN-NH3	P	30.9153 mg/l	11/12/2025 12:05:02
Q3530-01	S	TKN-NH3	P	0.2319 mg/l	11/12/2025 12:15:35
Q3530-02	S	TKN-NH3	P	0.5137 mg/l	11/12/2025 12:15:36
Q3586-02	S	TKN-NH3	P	19.7287 mg/l	11/12/2025 12:15:37
CCV4	S	TKN-NH3	P	5.118 mg/l	11/12/2025 12:15:38
CCB4	S	TKN-NH3	P	0.1611 mg/l	11/12/2025 12:15:39
Q3586-02DUP	S	TKN-NH3	P	19.6598 mg/l	11/12/2025 12:15:40
Q3586-02MS	S	TKN-NH3	P	22.2229 mg/l	11/12/2025 12:15:41
Q3586-02MSD	S	TKN-NH3	P	22.3735 mg/l	11/12/2025 12:15:42
Q3552-01DLX20	S	TKN-NH3	P	7.9015 mg/l	11/12/2025 12:15:43
Q3552-02DLX2	S	TKN-NH3	P	6.5321 mg/l	11/12/2025 12:15:45
Q3552-03DLX10	S	TKN-NH3	P	8.5336 mg/l	11/12/2025 12:15:46
Q3552-05DLX10	S	TKN-NH3	P	6.6612 mg/l	11/12/2025 12:26:19
Q3560-07DLX10	S	TKN-NH3	P	4.8544 mg/l	11/12/2025 12:26:20
Q3569-01DLX10	S	TKN-NH3	P	3.5942 mg/l	11/12/2025 12:26:21
CCV5	S	TKN-NH3	P	4.9652 mg/l	11/12/2025 13:42:24
CCB5	S	TKN-NH3	P	-0.0228 mg/l	11/12/2025 13:42:26
Q3569-02DLX10	S	TKN-NH3	P	7.0982 mg/l	11/12/2025 13:42:27
Q3483-07DLX10	S	TKN-NH3	P	5.586 mg/l	11/12/2025 13:42:28
Q3586-02DLX5	S	TKN-NH3	P	4.5261 mg/l	11/12/2025 13:42:31
Q3586-02DUPDLX5	S	TKN-NH3	P	4.4275 mg/l	11/12/2025 13:42:32
Q3484-07DLX20	S	TKN-NH3	P	6.876 mg/l	11/12/2025 13:42:33
CCV6	S	TKN-NH3	P	5.1821 mg/l	11/12/2025 13:42:34
CCB6	S	TKN-NH3	P	0.0153 mg/l	11/12/2025 13:45:39

Alliance Technical Group
 284 Sheffield Street, Mountainside, NJ 07092

11/12/2025 10:27 Reviewed by : RM Instrument ID : Konelab


Test TKN-NH3

Accepted 11/12/2025 10:27

Factor 5.126
 Bias 0.023

Coeff. of det. 0.998783

Errors



	Calibrator	Response	Calc. con.	Conc.	^{1e} Errors
1	0.00PPM	0.021	-0.0130	0.0000	-
2	TKN-10	0.112	0.4527	0.5000	-9.5
3	TKN-10	0.202	0.9186	1.0000	-8.1
4	TKN-10	0.513	2.5078	2.5000	0.3
5	TKN-10	1.020	5.1098	5.0000	2.2
6	TKN-10	1.365	6.8781	6.6667	2.7
7	TKN-10	1.937	9.8128	10.0000	-1.9

11/12/2025
 RM