

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137879

Slope : 98.3

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/13/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.01	11/13/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	11/13/2025	08:17
4	ICV	1	Water	NA	NA	20.2	7.01	11/13/2025	08:20
5	CCV1	1	Water	NA	NA	20.2	2.01	11/13/2025	08:22
6	Q3618-01	1	Solid	20.02	20	21.4	6.53	11/13/2025	08:35
7	Q3618-02	1	Solid	20.03	20	21.4	3.48	11/13/2025	08:45
8	Q3618-02DUP	1	Solid	20.04	20	21.6	3.50	11/13/2025	08:48
9	CCV2	1	Water	NA	NA	20.3	12.02	11/13/2025	08:50

WORKLIST(Hardcopy Internal Chain)

VB 131819

WorkList Name : ph s q3604 WorkList ID : 193091 Department : Wet-Chemistry Date : 11-13-2025 07:58:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3618-01	EME-TOP-SOIL	Solid	pH	Cool 4 deg C	ENTA05	D41	11/12/2025	9045D
Q3618-02	EME-GENERAL-FILL	Solid	pH	Cool 4 deg C	ENTA05	D41	11/12/2025	9045D

Date/Time 11-13-25 08:10
 Raw Sample Received by: JF WCC1
 Raw Sample Relinquished by: JF WCC1

Date/Time 11-13-25 13:00
 Raw Sample Received by: JF WCC1
 Raw Sample Relinquished by: JF WCC1