

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137880

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/13/2025	08:38
2	CAL2	1	Water	NA	NA	20.2	7.00	11/13/2025	08:39
3	CAL3	1	Water	NA	NA	20.3	10.02	11/13/2025	08:40
4	ICV	1	Water	NA	NA	20.1	7.00	11/13/2025	08:44
5	CCV1	1	Water	NA	NA	20.3	2.01	11/13/2025	08:45
6	Q3620-01	1	Water	NA	NA	20.4	5.26	11/13/2025	08:50
7	Q3620-02	1	Water	NA	NA	20.7	5.37	11/13/2025	09:00
8	Q3620-02DUP	1	Water	NA	NA	20.8	5.39	11/13/2025	09:05
9	CCV2	1	Water	NA	NA	20.2	12.02	11/13/2025	09:10

WORKLIST(Hardcopy Internal Chain)

UP 134880

WorkList Name : ph w q3622 WorkList ID : 193092 Department : Wet-Chemistry Date : 11-13-2025 07:59:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3620-01	1010-A	Water	pH	Cool 4 deg C	PSEG03	D31	11/12/2025	9040C
Q3620-02	1010-B	Water	pH	Cool 4 deg C	PSEG03	D31	11/12/2025	9040C

Date/Time 11-13-25 08:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 11-13-25 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]