

7470A


LB137926

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42555



A= 0.0000e+000

B= 2.3434e-004

C= -4.8898e-002

Rho= 0.9997702

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	66	0.000	66	0			
0.2	0.200	0.142	-0.058	816	0.0 %	816	0			
2.5	2.500	2.522	0.022	10972	0.0 %	10972	0			
5.0	5.000	5.163	0.163	22239	0.0 %	22239				
7.5	7.500	7.482	-0.018	32138	0.0 %	32138				
10.0	10.000	9.924	-0.076	42555	0.0 %	42555				

LB137920 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Nov 2025 14:38:51

Type	Sample ID	Extended ID	Conc	µ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	66	PPB	Std	0.0000	7470A	17 Nov 2025 14:41:54
S	0.2 - 1	50.2	-	816	PPB	Std	0.2000	7470A	17 Nov 2025 14:46:52
S	2.5 - 1	52.5	-	10972	PPB	Std	2.5000	7470A	17 Nov 2025 14:49:09
S	5.0 - 1	55	-	22239	PPB	Std	5.0000	7470A	17 Nov 2025 14:51:26
S	7.5 - 1	57.5	-	32138	PPB	Std	7.5000	7470A	17 Nov 2025 14:53:42
S	10.0 - 1	59	-	42555	PPB	Std	10.0000	7470A	17 Nov 2025 15:11:36
U	ICV118 - 1	ICV118	3.9495	17062	PPB	SMPL	-7470A		17 Nov 2025 15:17:04
U	ICB118 - 1	ICB118	-0.0524	-15	PPB	SMPL	-7470A		17 Nov 2025 15:19:29
U	CCV04 - 1	CCV04	5.0223	21640	PPB	SMPL	-7470A		17 Nov 2025 15:21:44
U	CCB04 - 1	CCB04	-0.0571	-35	PPB	SMPL	-7470A		17 Nov 2025 15:23:59
U	CRA - 1	CRA	0.1641	909	PPB	SMPL	-7470A		17 Nov 2025 15:26:14
U	HighStd - 1	HighStd	9.7999	42027	PPB	SMPL	-7470A		17 Nov 2025 15:28:37
U	ChkStd - 1	ChkStd	7.3334	31502	PPB	SMPL	-7470A		17 Nov 2025 15:30:52
U	PB170585BL - 1	PBW	-0.0721	-99	PPB	SMPL	-7470A		17 Nov 2025 15:33:10
U	PB170585BS - 1	LCSW	3.7615	16260	PPB	SMPL	-7470A		17 Nov 2025 15:35:27
U	Q3628-04 - 1	TP-5	-0.0791	-129	PPB	SMPL	-7470A		17 Nov 2025 15:37:43
U	Q3638-02 - 1	MOO-25-0327	-0.0576	-37	PPB	SMPL	-7470A		17 Nov 2025 15:40:00
U	Q3638-04 - 1	MOO-25-0328	-0.0451	16	PPB	SMPL	-7470A		17 Nov 2025 15:42:17
U	Q3640-02 - 1	LAW-25-0179-0180	-0.0372	50	PPB	SMPL	-7470A		17 Nov 2025 15:44:34
U	Q3640-03 - 1	LAW-25-0182	-0.0369	51	PPB	SMPL	-7470A		17 Nov 2025 15:46:49
U	CCV05 - 1	CCV05	4.9902	21503	PPB	SMPL	-7470A		17 Nov 2025 15:49:04
U	CCB05 - 1	CCB05	-0.0616	-54	PPB	SMPL	-7470A		17 Nov 2025 15:51:19
U	Q3647-02 - 1	VNJ-212	-0.0430	25	PPB	SMPL	-7470A		17 Nov 2025 15:53:35
U	Q3649-07 - 1	B4-0.0-0.5-20251114	-0.0421	29	PPB	SMPL	-7470A		17 Nov 2025 15:55:50
U	Q3649-07DUP - 1	B4-0.0-0.5-20251114DUP	-0.0421	29	PPB	SMPL	-7470A		17 Nov 2025 15:58:05
U	Q3649-07MS - 1	B4-0.0-0.5-20251114MS	3.8318	16560	PPB	SMPL	-7470A		17 Nov 2025 16:00:21
U	Q3649-07MSD - 1	B4-0.0-0.5-20251114MSD	3.7362	16152	PPB	SMPL	-7470A		17 Nov 2025 16:02:36
U	PB170544TB - 1	PB170544TB	-0.0866	-161	PPB	SMPL	-7470A		17 Nov 2025 16:04:53
U	Q3649-07LX5 - 1		-0.0536	-20	PPB	SMPL	-7470A		17 Nov 2025 16:07:10
U	Q3649-07A - 1		3.8443	16613	PPB	SMPL	-7470A		17 Nov 2025 16:09:27
U	CCV06 - 1	CCV06	4.9825	21470	PPB	SMPL	-7470A		17 Nov 2025 16:11:44
U	CCB06 - 1	CCB06	-0.0606	-50	PPB	SMPL	-7470A		17 Nov 2025 16:14:00