

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137937

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/18/2025	10:40
2	CAL2	1	Water	NA	NA	20.2	7.00	11/18/2025	10:41
3	CAL3	1	Water	NA	NA	20.2	10.02	11/18/2025	10:42
4	ICV	1	Water	NA	NA	20.2	6.98	11/18/2025	10:44
5	CCV1	1	Water	NA	NA	20.2	2.01	11/18/2025	10:45
6	Q3640-04	1	Water	NA	NA	20.9	5.30	11/18/2025	10:50
7	Q3660-01	1	Water	NA	NA	20.1	12.91	11/18/2025	11:10
8	Q3660-01DUP	1	Water	NA	NA	20.2	12.93	11/18/2025	11:11
9	CCV2	1	Water	NA	NA	20.3	12.02	11/18/2025	11:12

WORKLIST(Hardcopy Internal Chain)

87 137937

WorkList Name : ph q3660

WorkList ID : 193175

Department : Wet-Chemistry

Date : 11-18-2025 10:20:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3640-04	B LAW-25-0183	Water	pH	Cool 4 deg C	PSEG03	D41	11/13/2025	9040C
Q3660-01	L FRAC-TANK-M21012	Water	pH	Cool 4 deg C	PSEG03	D41	11/17/2025	9040C

Date/Time 11-18-25 10:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 11-18-25 13:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]