

Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137938

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.02	11/18/2025	11:30
2	CAL2	1	Water	NA	NA	20.3	7.01	11/18/2025	11:32
3	CAL3	1	Water	NA	NA	20.3	10.02	11/18/2025	11:33
4	ICV	1	Water	NA	NA	20.3	7.00	11/18/2025	11:35
5	CCV1	1	Water	NA	NA	20.3	2.01	11/18/2025	11:37
6	Q3652-04	1	Solid	20.02	20	21.1	6.59	11/18/2025	11:40
7	Q3652-04DUP	1	Solid	20.03	20	21.2	6.60	11/18/2025	11:42
8	Q3659-01	1	Solid	20.03	20	21.1	7.06	11/18/2025	11:50
9	Q3659-02	1	Solid	20.02	20	21.6	6.67	11/18/2025	12:00
10	CCV2	1	Water	NA	NA	20.3	12.02	11/18/2025	12:02

WORKLIST(Hardcopy Internal Chain)

W 121938

WorkList Name : corrosivity q3652

WorkList ID : 193177

Department : Wet-Chemistry

Date : 11-18-2025 10:23:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3652-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/17/2025	9045D
Q3659-01	TANK-1-COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG04	D41	11/17/2025	9045D
Q3659-02	TANK-1-COMP-2	Solid	Corrosivity	Cool 4 deg C	PSEG04	D41	11/17/2025	9045D

Date/Time 11/18/25 10:30
Raw Sample Received by: SA WDC
Raw Sample Relinquished by: [Signature]

Date/Time 11/18/25 14:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]