


L3137942

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3717e-004

C= -5.1121e-003

Rho= 0.9999511

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.014  | 0.014  | 80    | 0.000      | 80    | 0     |       |       |       |
| 0.2    | 0.200  | 0.189  | -0.011 | 818   | 0.0 %      | 818   | 0     |       |       |       |
| 2.5    | 2.500  | 2.532  | 0.032  | 10699 | 0.0 %      | 10699 | 0     |       |       |       |
| 5.0    | 5.000  | 4.978  | -0.022 | 21010 | 0.0 %      | 21010 |       |       |       |       |
| 7.5    | 7.500  | 7.438  | -0.062 | 31385 | 0.0 %      | 31385 |       |       |       |       |
| 10.0   | 10.000 | 10.049 | 0.049  | 42390 | 0.0 %      | 42390 |       |       |       |       |

## LB137942 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Nov 2025 13:31:18

| Type | Sample ID       | Extended ID         | Conc    | μ Abs | Units | Type | Std Conc | Method | Date                 |
|------|-----------------|---------------------|---------|-------|-------|------|----------|--------|----------------------|
| S    | 0.0 - 1         |                     | -       | 80    | PPB   | Std  | 0.0000   | 7470A  | 18 Nov 2025 13:49:00 |
| S    | 0.2 - 1         |                     | -       | 818   | PPB   | Std  | 0.2000   | 7470A  | 18 Nov 2025 13:54:38 |
| S    | 2.5 - 1         |                     | -       | 10699 | PPB   | Std  | 2.5000   | 7470A  | 18 Nov 2025 13:56:55 |
| S    | 5.0 - 1         |                     | -       | 21010 | PPB   | Std  | 5.0000   | 7470A  | 18 Nov 2025 13:59:12 |
| S    | 7.5 - 1         |                     | -       | 31385 | PPB   | Std  | 7.5000   | 7470A  | 18 Nov 2025 14:01:29 |
| S    | 10.0 - 1        |                     | -       | 42390 | PPB   | Std  | 10.0000  | 7470A  | 18 Nov 2025 14:03:45 |
| U    | ICV120 - 1      | ICV120              | 4.0685  | 17176 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:07:13 |
| U    | ICB120 - 1      | ICB120              | -0.0125 | -31   | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:09:27 |
| U    | CCV10 - 1       | CCV10               | 4.9228  | 20778 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:11:43 |
| U    | CCB10 - 1       | CCB10               | -0.0217 | -70   | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:13:57 |
| U    | CRA - 1         | CRA                 | 0.1562  | 680   | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:16:12 |
| U    | HighStd - 1     | HighStd             | 9.8458  | 41535 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:18:27 |
| U    | ChkStd - 1      | ChkStd              | 6.8086  | 28729 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:20:42 |
| U    | PB170602BL - 1  | PBW                 | -0.0284 | -98   | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:23:00 |
| U    | PB170602BS - 1  | LCSW                | 3.7861  | 15985 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:25:17 |
| U    | Q3660-01 - 1    | FRAC-TANK-M21012    | 2.0417  | 8630  | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:27:34 |
| U    | Q3660-01DUP - 1 | FRAC-TANK-M21012DUP | 1.9978  | 8445  | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:29:50 |
| U    | Q3660-01MS - 1  | FRAC-TANK-M21012MS  | 5.3151  | 22432 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:32:08 |
| U    | Q3660-01MSD - 1 | FRAC-TANK-M21012MSD | 5.2323  | 22083 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:34:26 |
| U    | CCV11 - 1       | CCV11               | 4.9202  | 20767 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:36:43 |
| U    | CCB11 - 1       | CCB11               | -0.0312 | -110  | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:38:00 |
| U    | Q3660-01LX5 - 1 |                     | 0.4488  | 1914  | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:41:15 |
| U    | Q3660-01A - 1   |                     | 5.7088  | 24092 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:43:30 |
| U    | CCV12 - 1       | CCV12               | 4.7478  | 20040 | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:48:11 |
| U    | CCB12 - 1       | CCB12               | -0.0205 | -65   | PPB   | SMPL | -7470A   |        | 18 Nov 2025 14:50:26 |