



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/19/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 11/18/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:35
Out Date: 11/19/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137951

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3662-01	B3-0.5-1.0-20251117	1	1.15	10.55	11.7	10.77	91.2	
Q3662-02	B3-0.0-1.0-20251117	2	1.19	10.67	11.86	10.23	84.7	
Q3662-03	B1-0.5-1.0-20251117	3	1.16	10.37	11.53	10.76	92.6	
Q3662-04	B1-0.0-1.0-20251117	4	1.17	10.62	11.79	10.53	88.1	
Q3663-01	2053	5	1.00	1.00	2.00	2.00	100.0	oil sample
Q3663-02	2054	6	1.00	1.00	2.00	2.00	100.0	oil sample
Q3664-01	ARS20-0001	7	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3664-02	ARS20-0001-E2	8	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3664-03	ARS20-0013	9	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3664-04	ARS20-0013-E2	10	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3664-05	291399	11	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE, 100% SOLIDS
Q3664-06	291399-E2	12	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE, 100% SOLIDS
Q3665-01	BUR-25-0101	13	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3666-01	BUR-25-0103	14	1.18	10.38	11.56	10.66	91.3	
Q3667-01	RIGHT-SIDE-WALL	15	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3667-03	LEFT-SIDE-WALL	16	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3667-05	LEFT-SIDE-FLOOR	17	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3667-07	ELEVATED-DIVIDE	18	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3668-01	2015	19	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3669-01	EO-01-11182025	20	1.11	10.43	11.54	10.8	92.9	
Q3669-02	EO-01-11182025-E2	21	1.15	9.98	11.13	10.33	92.0	
Q3669-02D	EO-01-11182025-E2DUP	22	1.19	9.96	11.15	10.35	92.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$