

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137966

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	11/19/2025	09:15
2	Q3574-01		1	9	88.00	190.4	190.4	11/19/2025	09:45
3	Q3574-03		1	9	92.00	197.6	197.6	11/19/2025	10:35
4	Q3660-01		1	14	100.00	>212.0	>212.0	11/19/2025	11:05
5	Q3660-01DUP		1	14	100.00	>212.0	>212.0	11/19/2025	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16137966

WorkList Name : FP-

WorkList ID : 193193

Department : Wet-Chemistry

Date : 11-18-2025 17:10:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3574-01	304641-01 Liquid	Water	Flash Point	Cool 4 deg C	PACI01	J22	11/04/2025	1010B
Q3574-03	304614-01 fuel	Water	Flash Point	Cool 4 deg C	PACI01	J22	11/04/2025	1010B
Q3660-01	FRAC-TANK-M21012	Water	Flash Point	Cool 4 deg C	PSEG03	D41	11/17/2025	1010B

Date/Time 11/19/2025 08:45
Raw Sample Received by: RIM
Raw Sample Relinquished by: [Signature]

Date/Time 11/19/2025 12:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]