

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137973

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	11/20/2025	08:25
2	Q3663-06		1	13	100.00	>212.0	>212.0	11/20/2025	12:10
3	Q3663-06DUP		1	13	100.00	>212.0	>212.0	11/20/2025	12:40
4	Q3663-07		1	15	100.00	>212.0	>212.0	11/20/2025	13:10
5	Q3663-08		1	15	100.00	>212.0	>212.0	11/20/2025	13:40
6	Q3663-09		1	14	100.00	>212.0	>212.0	11/20/2025	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137973

WorkList Name : flash point water WorkList ID : 193224 Department : Wet-Chemistry Date : 11-19-2025 15:54:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3663-06	2055-56-57	Water	Flash Point	Cool 4 deg C	PSEG03	D31	11/18/2025	1010B
Q3663-07	60491	Water	Flash Point	Cool 4 deg C	PSEG03	D31	11/18/2025	1010B
Q3663-08	1102	Water	Flash Point	Cool 4 deg C	PSEG03	D31	11/18/2025	1010B
Q3663-09	1103	Water	Flash Point	Cool 4 deg C	PSEG03	D31	11/18/2025	1010B

Date/Time 11/20/2025 08:15
 Raw Sample Received by: RM Wey
 Raw Sample Relinquished by: JH(WC)

Date/Time 11/20/2025 14:50
 Raw Sample Received by: JH(WC)
 Raw Sample Relinquished by: RM Wey