

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137976

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.02	11/20/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	11/20/2025	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	11/20/2025	09:05
4	ICV	1	Water	NA	NA	20.3	7.02	11/20/2025	09:10
5	CCV1	1	Water	NA	NA	20.3	2.01	11/20/2025	09:11
6	Q3682-01	1	Water	NA	NA	20.8	7.52	11/20/2025	09:20
7	Q3682-01DUP	1	Water	NA	NA	20.9	7.53	11/20/2025	09:21
8	Q3685-01	1	Water	NA	NA	20.7	5.00	11/20/2025	09:30
9	Q3685-01DUP	1	Water	NA	NA	20.9	5.02	11/20/2025	09:33
10	Q3685-02	1	Water	NA	NA	20.8	5.24	11/20/2025	09:37
11	Q3685-02DUP	1	Water	NA	NA	20.9	5.26	11/20/2025	09:39
12	Q3685-03	1	Water	NA	NA	20.9	6.25	11/20/2025	09:45
13	Q3685-03DUP	1	Water	NA	NA	20.8	6.27	11/20/2025	09:47
14	Q3687-01	1	Water	NA	NA	20.4	6.71	11/20/2025	09:50
15	Q3687-01DUP	1	Water	NA	NA	20.6	6.72	11/20/2025	09:53
16	CCV2	1	Water	NA	NA	20.2	12.02	11/20/2025	09:55

WORKLIST(Hardcopy Internal Chain)

137976

WorkList Name : ph w q3687 WorkList ID : 193227 Department : Wet-Chemistry Date : 11-20-2025 07:46:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3682-01	B 3456	Water	pH	Cool 4 deg C	PSEG03	D41	11/19/2025	9040C
Q3685-01	C 4197	Water	pH	Cool 4 deg C	PSEG03	D11	11/19/2025	9040C
Q3685-02	C 4198	Water	pH	Cool 4 deg C	PSEG03	D11	11/19/2025	9040C
Q3685-03	B 4206	Water	pH	Cool 4 deg C	PSEG03	D11	11/19/2025	9040C
Q3687-01	A N43502	Water	pH	Cool 4 deg C	PSEG03	D41	11/19/2025	9040C

Date/Time 11-20-25 07:55

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

11-20-25

Raw Sample Received by:

[Signature]

Raw Sample Relinquished by:

[Signature]