


LB137986

7471B

INSTRUMENT ID: CV1

μ Abs.:

53894



Linear

A= 0.0000e+000

B= 1.8158e-004

C= 3.1567e-002

Rho= 0.9990532

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.043	0.043	65	0.000	65				
0.2	0.200	0.175	-0.025	788	0.0 %	788				
2.5	2.500	2.477	-0.023	13470	0.0 %	13470				
5.0	5.000	4.869	-0.131	26638	0.0 %	26638				
7.5	7.500	7.818	0.318	42881	0.0 %	42881				
10.0	10.000	9.818	-0.182	53894	0.0 %	53894				

LB137986 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Nov 2025 11:03:55

Type	Sample ID	Extended ID	Conc	μ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	CG-6	-	65	PPB	Std	0.0000	7471B	20 Nov 2025 11:52:50
S	0.2 - 1	CG-2	-	788	PPB	Std	0.2000	7471B	20 Nov 2025 11:55:16
S	2.5 - 1		-	13470	PPB	Std	2.5000	7471B	20 Nov 2025 11:57:33
S	5.0 - 1		-	26638	PPB	Std	5.0000	7471B	20 Nov 2025 12:02:39
S	7.5 - 1		-	42881	PPB	Std	7.5000	7471B	20 Nov 2025 12:04:55
S	10.0 - 1		-	53894	PPB	Std	10.0000	7471B	20 Nov 2025 12:07:14
U	ICV123 - 1	ICV123	4.2171	23050	PPB	SMPL	-7471B		20 Nov 2025 12:11:42
U	ICB123 - 1	ICB123	0.0152	-90	PPB	SMPL	-7471B		20 Nov 2025 12:13:57
U	CCV19 - 1	CCV19	4.6452	25408	PPB	SMPL	-7471B		20 Nov 2025 12:16:12
U	CCB19 - 1	CCB19	-0.0100	-229	PPB	SMPL	-7471B		20 Nov 2025 12:18:27
U	CRA - 1	CRA	0.1723	775	PPB	SMPL	-7471B		20 Nov 2025 12:20:45
U	HighStd - 1	HighStd	9.7790	53680	PPB	SMPL	-7471B		20 Nov 2025 12:23:00
U	ChkStd - 1	ChkStd	7.3470	40287	PPB	SMPL	-7471B		20 Nov 2025 12:25:15
U	PB170655BL - 1	PBS	-0.0126	-243	PPB	SMPL	-7471B		20 Nov 2025 12:27:33
U	PB170655BS - 1	LCSS	4.0197	21963	PPB	SMPL	-7471B		20 Nov 2025 12:29:51
U	Q3671-01 - 1	Zeitmann Power Plant HSRG 7B	0.2478	1191	PPB	SMPL	-7471B		20 Nov 2025 12:32:07
U	Q3678-01 - 1	COMPOSITE-STONE	0.0472	86	PPB	SMPL	-7471B		20 Nov 2025 12:34:26
U	Q3681-01 - 1	ETGI-354	0.9317	4957	PPB	SMPL	-7471B		20 Nov 2025 12:42:34
U	Q3681-01DUP - 1	ETGI-354DUP	0.9128	4853	PPB	SMPL	-7471B		20 Nov 2025 12:44:51
U	Q3681-01MS - 1	ETGI-354MS	4.5014	24616	PPB	SMPL	-7471B		20 Nov 2025 12:47:06
U	CCV20 - 1	CCV20	4.8386	26473	PPB	SMPL	-7471B		20 Nov 2025 12:49:22
U	CCB20 - 1	CCB20	0.0027	-159	PPB	SMPL	-7471B		20 Nov 2025 12:51:40
U	Q3681-01MSD - 1	ETGI-354MSD	4.2554	23261	PPB	SMPL	-7471B		20 Nov 2025 12:53:57
U	Q3686-01 - 1	FAIRLAWN-GRAVEL	0.0167	-82	PPB	SMPL	-7471B		20 Nov 2025 12:56:21
U	Q3681-01LX5 - 1		0.2237	1058	PPB	SMPL	-7471B		20 Nov 2025 12:58:38
U	CCV21 - 1	CCV21	4.3186	23609	PPB	SMPL	-7471B		20 Nov 2025 13:00:54
U	CCB21 - 1	CCB21	5.4912	30067	PPB	SMPL	-7471B		20 Nov 2025 13:06:18
			0.0091	-124	PPB	SMPL	-7471B		20 Nov 2025 13:08:35