

66137995

Test results

Aquakem 7.2AQ1

Page: 1

Alliance Technical Group

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

11/20/2025 14:21

Test: CNEPA-NEW

Sample Id	Result	Dil. 1 +	Response	Errors
ICV001 ICV001	96.220	0.0	0.095	
ICB001 ICB001	1.169	0.0	0.008	
CCV001 CCV001	242.903	0.0	0.229	
CCB001 CCB001	0.999	0.0	0.008	
PB170664BL PBS664	0.553	0.0	0.007	
Q3589-03 -MDL-SOIL-QT4-2025-07	4.544	0.0	0.011	
PB170665BL PBW665	0.278	0.0	0.007	
Q3589-01- MDL- WATER-QT4-2025-07	4.795	0.0	0.011	
11/20/2025 CCV002 CCV002	250.385	0.0	0.235	
RM CCB002 CCB002	0.312	0.0	0.007	

N 10
Mean 60.216
SD 102.5994
CV% 170.39

Aquakem v. 7.2AQ1

Results from time period:

Thu Nov 20 13:02:23 2025

Thu Nov 20 14:18:54 2025

Sample Id	Sam	Test short r	Test Result	Result unit	Result date and time
S0.0	A	CNEPA-NE\ P	0.5968	µg/l	11/20/2025 13:02:23
S5.0	A	CNEPA-NE\ P	4.6324	µg/l	11/20/2025 13:02:24
S10.0	A	CNEPA-NE\ P	9.8483	µg/l	11/20/2025 13:02:25
S100.0	A	CNEPA-NE\ P	100.1232	µg/l	11/20/2025 13:02:26
S250.0	A	CNEPA-NE\ P	249.6346	µg/l	11/20/2025 13:02:27
S500.0	A	CNEPA-NE\ P	500.1648	µg/l	11/20/2025 13:02:28
ICV001 ICV001	S	CNEPA-NE\ P	96.2199	µg/l	11/20/2025 13:38:20
ICB001 ICB001	S	CNEPA-NE\ P	1.1693	µg/l	11/20/2025 13:38:22
CCV001 CCV001	S	CNEPA-NE\ P	242.9034	µg/l	11/20/2025 13:38:24
CCB001 CCB001	S	CNEPA-NE\ P	0.9986	µg/l	11/20/2025 13:38:27
PB170664BL PBS664	S	CNEPA-NE\ P	0.5528	µg/l	11/20/2025 13:38:28
Q3589-03-MDL-SOIL-QT4-2025-07	S	CNEPA-NE\ P	4.5442	µg/l	11/20/2025 13:38:30
PB170665BL PBW665	S	CNEPA-NE\ P	0.2783	µg/l	11/20/2025 13:45:01
Q3589-01-MDL-WATER-QT4-2025-07	S	CNEPA-NE\ P	4.7952	µg/l	11/20/2025 14:18:50
CCV002 CCV002	S	CNEPA-NE\ P	250.3845	µg/l	11/20/2025 14:18:51
CCB002 CCB002	S	CNEPA-NE\ P	0.3123	µg/l	11/20/2025 14:18:54

Alliance Technical Group
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

11/20/2025 13:02


Test CNEPA-NEW

Accepted 11/20/2025 13:02

Factor slope 1095 0.000913
Bias y. intercept 0.007
11/20/2025 RM

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.0PPBCN\$0.0	0.007	0.5968	0.0000	-7.4
2	5.0PPBCN\$5.0	0.011	4.6324	5.0000	-1.5
3	10PPBCN\$10.0	0.016	9.8483	10.0000	0.1
4	100PPBCN\$100.0	0.098	100.1232	100.0000	-0.1
5	250PPBCN\$250.0	0.235	249.6346	250.0000	0.0
6	500PPBCN\$500.0	0.464	500.1648	500.0000	

11/20/2025
RM