

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB138000

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/21/2025	10:10
2	CAL2	1	Water	NA	NA	20.2	7.00	11/21/2025	10:11
3	CAL3	1	Water	NA	NA	20.3	10.02	11/21/2025	10:12
4	ICV	1	Water	NA	NA	20.3	7.02	11/21/2025	10:15
5	CCV1	1	Water	NA	NA	20.3	2.01	11/21/2025	10:18
6	Q3697-01	1	Solid	20.02	20	20.8	7.41	11/21/2025	10:25
7	Q3697-01DUP	1	Solid	20.03	20	20.9	7.42	11/21/2025	10:26
8	Q3697-02	1	Solid	20.02	20	22.1	6.52	11/21/2025	10:35
9	Q3697-03	1	Solid	20.03	20	20.8	8.50	11/21/2025	10:38
10	CCV2	1	Water	NA	NA	20.2	12.02	11/21/2025	10:40

WORKLIST(Hardcopy Internal Chain)

18 ~~1338000~~ 138000

WorkList Name : corrosivity q3697 WorkList ID : 193283 Department : Wet-Chemistry Date : 11-21-2025 09:50:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3697-01	MOO-25-0331	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/20/2025	9045D
Q3697-02	MOO-25-0332	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/20/2025	9045D
Q3697-03	MOO-25-0333	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	11/20/2025	9045D

Date/Time 11-21-25 10:00
 Raw Sample Received by: YB WDC
 Raw Sample Relinquished by: AS

Date/Time 11-21-25 13:30
 Raw Sample Received by: CP
 Raw Sample Relinquished by: YB WDC