

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138001

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/21/2025	10:37
2	CAL2	1	Water	NA	NA	20.3	6.99	11/21/2025	10:38
3	CAL3	1	Water	NA	NA	20.3	10.02	11/21/2025	10:40
4	ICV	1	Water	NA	NA	20.2	7.01	11/21/2025	10:44
5	CCV1	1	Water	NA	NA	20.3	2.01	11/21/2025	10:45
6	Q3697-04	1	Water	NA	NA	20.3	7.92	11/21/2025	10:50
7	Q3697-04DUP	1	Water	NA	NA	20.4	7.94	11/21/2025	10:52
8	CCV2	1	Water	NA	NA	20.2	12.02	11/21/2025	10:55

WORKLIST(Hardcopy Internal Chain)

VB 13801

WorkList Name : ph w q3697 WorkList ID : 193282 Department : Wet-Chemistry Date : 11-21-2025 09:49:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3697-04 B	MOO-25-0329	Water	pH	Cool 4 deg C	PSEG03	D41	11/20/2025	9040C

Date/Time 11.21.25 09:15:00
Raw Sample Received by: JB WDC1
Raw Sample Relinquished by: AP SR

Date/Time 11.21.25 12:13:5
Raw Sample Received by: AP SR
Raw Sample Relinquished by: JB WDC1