

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138006

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/21/2025	08:30
2	Q3666-02		1	15	100.00	>212.0	>212.0	11/21/2025	09:00
3	Q3666-02DUP		1	15	100.00	>212.0	>212.0	11/21/2025	09:30
4	Q3682-01		1	13	100.00	>212.0	>212.0	11/21/2025	10:00
5	Q3685-01		1	15	100.00	>212.0	>212.0	11/21/2025	10:30
6	Q3685-02		1	14	100.00	>212.0	>212.0	11/21/2025	11:00
7	Q3685-03		1	12	100.00	>212.0	>212.0	11/21/2025	11:30
8	Q3687-01		1	14	100.00	>212.0	>212.0	11/21/2025	12:00
9	Q3697-04		1	14	100.00	>212.0	>212.0	11/21/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06