


LB138008
INSTRUMENT ID: CV,

7470A

μ Abs.:

46188



Linear

A= 0.0000e+000

B= 2.1780e-004

C= -1.7129e-002

Rho= 0.9987042

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	6	0.000	6	0			
0.2	0.200	0.242	0.042	1189	0.0 %	1189	0			
2.5	2.500	2.356	-0.144	10894	0.0 %	10894	0			
5.0	5.000	5.341	0.341	24603	0.0 %	24603				
7.5	7.500	7.235	-0.265	33296	0.0 %	33296				
10.0	10.000	10.042	0.042	46188	0.0 %	46188				

LB138008 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Nov 2025 14:00:35

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	5.0	-	6	PPB	Std	0.0000	7470A	21 Nov 2025 14:04:14
S	0.2 - 1	50.2	-	1189	PPB	Std	0.2000	7470A	21 Nov 2025 14:06:32
S	2.5 - 1	52.5	-	10894	PPB	Std	2.5000	7470A	21 Nov 2025 14:08:50
S	5.0 - 1	55.0	-	24603	PPB	Std	5.0000	7470A	21 Nov 2025 14:11:06
S	7.5 - 1	57.5	-	33296	PPB	Std	7.5000	7470A	21 Nov 2025 14:13:23
S	10.0 - 1	59.0	-	46188	PPB	Std	10.0000	7470A	21 Nov 2025 14:15:38
U	ICV125 - 1	ICV125	3.8690	17843	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:19:51
U	ICB125 - 1	ICB125	-0.0518	-159	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:22:06
U	CCV24 - 1	CCV24	5.2684	24268	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:24:21
U	CCB24 - 1	CCB24	-0.0478	-141	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:26:36
U	CRA - 1	CRA	0.2342	1154	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:28:52
U	HighStd - 1	HighStd	9.8617	45358	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:31:07
U	ChkStd - 1	ChkStd	6.6875	30784	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:33:35
U	PB170676BL - 1	PBW	-0.0396	-103	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:35:53
U	PB170676BS - 1	LCSW	3.9378	18159	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:38:10
U	Q3682-01 - 1	3456	-0.0030	65	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:40:26
U	Q3685-01 - 1	4197	0.0393	259	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:42:43
U	Q3685-02 - 1	4198	0.2216	1096	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:45:59
U	Q3685-03 - 1	4206	0.0160	152	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:48:16
U	Q3687-01 - 1	N43502	0.1499	767	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:50:32
U	CCV25 - 1	CCV25	5.3897	24825	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:52:47
U	CCB25 - 1	CCB25	-0.0298	-58	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:55:02
U	Q3687-01DUP - 1	N43502DUP	0.1414	728	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:57:18
U	Q3687-01MS - 1	N43502MS	3.8895	17937	PPB	SMPL	-7470A	-7470A	21 Nov 2025 14:59:33
U	Q3687-01MSD - 1	N43502MSD	3.3515	15467	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:01:49
U	Q3697-04 - 1	MOO-25-0329	0.0733	415	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:04:05
U	PB170684BL - 1	PBW	-0.0354	-84	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:06:21
U	PB170684BS - 1	LCSW	3.4890	16098	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:08:37
U	Q3678-02 - 1	COMPOSITE-STONE	-0.0420	-114	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:10:53
U	Q3697-01 - 1	MOO-25-0331	-0.0197	-12	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:13:11
U	Q3697-02 - 1	MOO-25-0332	-0.0069	47	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:15:28
U	Q3697-03 - 1	MOO-25-0333	-0.0028	66	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:17:44
U	CCV26 - 1	CCV26	5.1630	23784	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:23:16
U	CCB26 - 1	CCB26	-0.0326	-71	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:25:31
U	Q3697-03DUP - 1	MOO-25-0333DUP	-0.0097	34	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:27:47
U	Q3697-03MS - 1	MOO-25-0333MS	3.6299	16745	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:30:03
U	Q3697-03MSD - 1	MOO-25-0333MSD	3.4630	15979	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:32:18
U	PB170645TB - 1	Q3697-03	-0.0437	-122	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:34:34
U	Q3687-01LX5 - 1		-0.0041	60	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:40:00
U	Q3687-01A - 1		3.8597	17800	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:42:16
U	Q3697-03LX5 - 1		-0.0008	75	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:48:09
U	Q3697-03A - 1		4.2127	19421	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:50:26
U	CCV27 - 1	CCV27	4.6222	21301	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:52:52
U	CCB27 - 1	CCB27	-0.0437	-122	PPB	SMPL	-7470A	-7470A	21 Nov 2025 15:55:10