

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB138047

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/26/2025	12:20
2	CAL2	1	Water	NA	NA	20.2	7.00	11/26/2025	12:22
3	CAL3	1	Water	NA	NA	20.3	10.02	11/26/2025	12:23
4	ICV	1	Water	NA	NA	20.3	7.01	11/26/2025	12:25
5	CCV1	1	Water	NA	NA	20.2	2.02	11/26/2025	12:28
6	Q3722-01	1	Water	NA	NA	20.1	6.70	11/26/2025	12:37
7	Q3722-01DUP	1	Water	NA	NA	20.2	6.73	11/26/2025	12:39
8	Q3722-02	1	Water	NA	NA	20.2	6.97	11/26/2025	12:45
9	Q3722-02DUP	1	Water	NA	NA	20.4	7.00	11/26/2025	12:48
10	CCV2	1	Water	NA	NA	20.3	12.02	11/26/2025	12:50

WORKLIST(Hardcopy Internal Chain)

VB 138048

WorkList Name : phw q3722

WorkList ID : 193361

Department : Wet-Chemistry

Date : 11-26-2025 12:06:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3722-01	257400	Water	pH	Cool 4 deg C	PSEG03	E12	11/25/2025	9040C
Q3722-02	259319	Water	pH	Cool 4 deg C	PSEG03	E12	11/25/2025	9040C

Date/Time 11-26-25 12:15
 Raw Sample Received by: JQ WOC
 Raw Sample Relinquished by: RJ (EIA-1066)

Date/Time

Raw Sample Received by:

11-26-25 12:30

Raw Sample Received by:

Raw Sample Relinquished by:

RJ (EIA-1066)

JQ WOC