

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138056

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	11/26/2025	11:15
2	Q3722-01		1	14	100.00	>212.0	>212.0	11/26/2025	11:45
3	Q3722-01DUP		1	14	100.00	>212.0	>212.0	11/26/2025	12:35
4	Q3722-02		1	15	100.00	>212.0	>212.0	11/26/2025	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6138056

WorkList Name : fp-3722

WorkList ID : 193370

Department : Wet-Chemistry

Date : 11-26-2025 09:47:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3722-01	257400	Water	Flash Point	Cool 4 deg C	PSEG03	E12	11/25/2025	1010B
Q3722-02	259319	Water	Flash Point	Cool 4 deg C	PSEG03	E12	11/25/2025	1010B

Date/Time 11/26/2025 10:50
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: RM (WC)

Date/Time 11/26/2025 15:40
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: RM (WC)