

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138074

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/02/2025	08:14
2	CAL2	1	Water	NA	NA	20.2	7.00	12/02/2025	08:15
3	CAL3	1	Water	NA	NA	20.3	10.02	12/02/2025	08:19
4	ICV	1	Water	NA	NA	20.2	7.01	12/02/2025	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	12/02/2025	08:30
6	Q3749-01	1	Water	NA	NA	20.8	8.51	12/02/2025	08:47
7	Q3749-01DUP	1	Water	NA	NA	20.9	8.53	12/02/2025	08:50
8	CCV2	1	Water	NA	NA	20.3	12.02	12/02/2025	08:55

WORKLIST(Hardcopy Internal Chain)

178074

WorkList Name : PH W Q3749 WorkList ID : 193405 Department : Wet-Chemistry Date : 12-02-2025 07:38:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3749-01	AUD-25-0204	Water	pH	Cool 4 deg C	PSEG05	A11	12/01/2025	9040C

Date/Time 12/02/25 07:46

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 12/02/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]