

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138075

Slope : 99.2

BalanceID: MSC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.02	12/02/2025	08:55
2	CAL2	1	Water	NA	NA	20.3	7.01	12/02/2025	08:56
3	CAL3	1	Water	NA	NA	20.3	10.02	12/02/2025	08:59
4	ICV	1	Water	NA	NA	20.3	7.01	12/02/2025	09:00
5	CCV1	1	Water	NA	NA	20.3	2.01	12/02/2025	09:10
6	Q3748-01	1	Solid	20.02	20	20.2	8.52	12/02/2025	09:20
7	Q3748-01DUP	1	Solid	20.03	20	20.3	8.54	12/02/2025	09:22
8	CCV2	1	Water	NA	NA	20.2	12.02	12/02/2025	09:25

WORKLIST(Hardcopy Internal Chain)

WP 138075

WorkList Name : PH S Q3748 WorkList ID : 193406 Department : Wet-Chemistry Date : 12-02-2025 07:38:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3748-01	AU-05-12-1-2025	Solid	pH	Cool 4 deg C	PSEG05	A12	12/01/2025	9045D

Date/Time 12/02/25 07:47
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12/02/25 11:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]