

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB138080

Slope : 98.5

BalanceID: MSC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/02/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/02/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/02/2025	08:12
4	ICV	1	Water	NA	NA	20.1	6.98	12/02/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.02	12/02/2025	08:16
6	Q3750-07	1	Solid	20.02	20	20.2	6.90	12/02/2025	08:25
7	Q3750-07DUP	1	Solid	20.03	20	20.4	6.92	12/02/2025	08:30
8	Q3750-08	1	Solid	20.05	20	20.2	6.99	12/02/2025	08:37
9	CCV2	1	Water	NA	NA	20.3	12.02	12/02/2025	08:40

WORKLIST(Hardcopy Internal Chain)

NY 138080

WorkList Name : corrosivity q3750 WorkList ID : 193407 Department : Wet-Chemistry Date : 12-02-2025 07:40:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3750-07	FENCE WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG05	A11	12/01/2025	9045D
Q3750-08	FENCE WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG05	A11	12/01/2025	9045D

Date/Time 12/02/25 07:55
Raw Sample Received by: JAWC
Raw Sample Relinquished by: JAWC

Date/Time 12/02/25 12:30
Raw Sample Received by: JAWC
Raw Sample Relinquished by: JAWC