

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138083

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	12/02/2025	14:40
2	Q3749-01		1	14	100.00	>212.0	>212.0	12/02/2025	15:10
3	Q3749-01DUP		1	14	100.00	>212.0	>212.0	12/02/2025	15:40
4	Q3755-01		1	13	100.00	>212.0	>212.0	12/02/2025	16:10
5	Q3755-02		1	14	100.00	>212.0	>212.0	12/02/2025	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16138083

WorkList Name : fp-12-02

WorkList ID : 193429

Department : Wet-Chemistry

Date : 12-02-2025 08:42:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3749-01	AUD-25-0204	Water	Flash Point	Cool 4 deg C	PSEG05	A11	12/01/2025	1010B
Q3755-01	TOTE#1108	Water	Flash Point	Cool 4 deg C	PSEG03	A11	12/02/2025	1010B
Q3755-02	#1114	Water	Flash Point	Cool 4 deg C	PSEG03	A11	12/02/2025	1010B

Date/Time 12/02/2025 14:30
 Raw Sample Received by: RH Cwr
 Raw Sample Relinquished by: CP 8m

Date/Time 12/02/2025 17:15
 Raw Sample Received by: CP 8m
 Raw Sample Relinquished by: RH Cwr