

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138087

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/03/2025	10:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/03/2025	10:11
3	CAL3	1	Water	NA	NA	20.3	10.03	12/03/2025	10:12
4	ICV	1	Water	NA	NA	20.3	7.02	12/03/2025	10:14
5	CCV1	1	Water	NA	NA	20.3	2.01	12/03/2025	10:15
6	Q3755-01	1	Water	NA	NA	20.1	5.45	12/03/2025	10:30
7	Q3755-02	1	Water	NA	NA	20.4	5.47	12/03/2025	10:37
8	Q3755-02DUP	1	Water	NA	NA	20.6	5.49	12/03/2025	10:39
9	CCV2	1	Water	NA	NA	20.2	12.02	12/03/2025	10:40

WORKLIST(Hardcopy Internal Chain)

JP 138087

WorkList Name : ph w q3755 WorkList ID : 193440 Department : Wet-Chemistry Date : 12-03-2025 09:41:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3755-01	☒ TOTE#1108	Water	pH	Cool 4 deg C	PSEG03	A11	12/02/2025	9040C
Q3755-02	☒ #1114	Water	pH	Cool 4 deg C	PSEG03	A11	12/02/2025	9040C

Date/Time 12/03/25 10:00
Raw Sample Received by: JP WCC
Raw Sample Relinquished by: JP WCC

Date/Time 12/03/25 12:50
Raw Sample Received by: JP WCC
Raw Sample Relinquished by: JP WCC