

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB138119

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP115951
Residual chlorine ICV-LCS, 0.4PPM	WP115950
Chlorine Calibration std, 0.1ppm	WP115945
Chlorine Calibration std, 0.2ppm	WP115946
Chlorine Calibration std, 0.8ppm	WP115947
Chlorine Calibration std, 0.0ppm	WP115944
Chlorine Calibration std, 1.6ppm	WP115948
Residual Chlorine Calibration and CCV std, 0	WP115949
Total Chlorine Powder Pillows	W3147

Intercept: 0.0054

Slope: 0.9895

Regression: 0.999774

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		12/04/2025	16:50
2	CAL2	0.1	1	0.000	0.110	0.110	0.11	6	12/04/2025	16:53
3	CAL3	0.2	1	0.000	0.200	0.200	0.20	-1.5	12/04/2025	16:56
4	CAL4	0.4	1	0.000	0.390	0.390	0.39	-2.8	12/04/2025	17:00
5	CAL5	0.8	1	0.000	0.820	0.820	0.82	2.9	12/04/2025	17:03
6	CAL6	1.6	1	0.000	1.580	1.580	1.59	-0.6	12/04/2025	17:06

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB138119

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.3860	12/04/2025	17:09
2	ICB				1	0.0000	0.0010	0.0010	-0.0410	12/04/2025	17:13
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.3970	12/04/2025	17:16
4	CCB1				1	0.0000	0.0010	0.0010	-0.0410	12/04/2025	17:20
5	LB138119BL	50	50		1	0.0000	0.0010	0.0010	-0.0410	12/04/2025	17:23
6	LB138119BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4190	12/04/2025	17:26
7	Q3530-09	50	50		1	0.0000	0.0600	0.0600	0.0550	12/04/2025	17:29
8	Q3785-04	50	50		1	0.0000	0.6600	0.6600	0.6470	12/04/2025	17:32
9	Q3785-04DUP	50	50		1	0.0000	0.6500	0.6500	0.6370	12/04/2025	17:35
10	Q3785-04MS	50	50	0.4	1	0.0000	1.0100	1.0100	1.0120	12/04/2025	17:38
11	Q3785-04MSD	50	50	0.4	1	0.0000	1.0200	1.0200	1.0230	12/04/2025	17:42
12	CCV2			0.4	1	0.0000	0.4100	0.4100	0.3860	12/04/2025	17:45
13	CCB2				1	0.0000	0.0010	0.0010	-0.0410	12/04/2025	17:49