

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138123

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	12/05/2025	12:25
2	Q3758-02		1	12	100.00	>212.0	>212.0	12/05/2025	12:55
3	Q3758-02DUP		1	12	100.00	>212.0	>212.0	12/05/2025	13:25
4	Q3759-01		1	14	100.00	>212.0	>212.0	12/05/2025	13:55
5	Q3775-01		1	9	61.00	141.8	141.5	12/05/2025	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66138123

WorkList Name : FP-12-03

WorkList ID : 193455

Department : Wet-Chemistry

Date : 12-03-2025 16:33:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3758-02	3462	Solid	Flash Point	Cool 4 deg C	PSEG03	A11	12/03/2025	1010B
Q3759-01	50851-50851B	Solid	Flash Point	Cool 4 deg C	PSEG03	A11	12/03/2025	1010B
Q3775-01	NGKO	Solid	Flash Point	Cool 4 deg C	PSEG04	--Sele	12/03/2025	1010B

Date/Time 12/15/2025 11:45
Raw Sample Received by: RY
Raw Sample Relinquished by: SP (W)

Date/Time 12/15/2025 15:10
Raw Sample Received by: SP (W)
Raw Sample Relinquished by: RY (W)