

L3138142


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

41302



A= 0.0000e+000

B= 2.4643e-004

C= 5.2875e-002

Rho= 0.9989157

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.064	0.064	44	0.000	44	0			
0.2	0.200	0.232	0.032	726	0.0 %	726	0			
2.5	2.500	2.627	0.127	10444	0.0 %	10444	0			
5.0	5.000	4.765	-0.235	19120	0.0 %	19120				
7.5	7.500	7.283	-0.217	29338	0.0 %	29338				
10.0	10.000	10.231	0.231	41302	0.0 %	41302				

LB138142 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Dec 2025 11:17:31

Type	Sample ID	Extended ID	Conc.	μ Abs	Units	Method	Std Conc	Date
S	0.0 - 1	5.0	-	44	PPB	7470A	0.0000	08 Dec 2025 12:14:38
S	0.2 - 1	50.2	-	726	PPB	7470A	0.2000	08 Dec 2025 12:16:55
S	2.5 - 1	500.5	-	10444	PPB	7470A	2.5000	08 Dec 2025 12:19:13
S	5.0 - 1	5.0	-	19120	PPB	7470A	5.0000	08 Dec 2025 12:21:30
S	7.5 - 1	50.0	-	29338	PPB	7470A	7.5000	08 Dec 2025 12:23:47
S	10.0 - 1	510	-	41302	PPB	7470A	10.0000	08 Dec 2025 12:29:56
U	ICV142 - 1	ICV142	4.0576	16251	PPB	7470A		08 Dec 2025 12:33:26
U	ICB142 - 1	ICB142	0.0307	-90	PPB	7470A		08 Dec 2025 12:35:41
U	CCV74 - 1	CCV74	4.7761	19167	PPB	7470A		08 Dec 2025 12:37:56
U	CCB74 - 1	CCB74	0.0305	-91	PPB	7470A		08 Dec 2025 12:40:12
U	CRA - 1	CRA	0.2084	631	PPB	7470A		08 Dec 2025 12:42:27
U	HighStd - 1	HighStd	9.4928	38307	PPB	7470A		08 Dec 2025 12:47:44
U	ChkStd - 1	ChkStd	7.2158	29067	PPB	7470A		08 Dec 2025 12:49:59
U	PB170845BL - 1	PBW	0.0144	-156	PPB	7470A		08 Dec 2025 12:52:17
U	PB170845BS - 1	LCSW	4.0221	16107	PPB	7470A		08 Dec 2025 12:54:35
U	Q3758-02 - 1	3462	0.0226	-123	PPB	7470A		08 Dec 2025 12:56:51
U	Q3770-01 - 1	4207	0.0093	-177	PPB	7470A		08 Dec 2025 12:59:09
U	Q3770-01DUP - 1	4207DUP	0.0371	-64	PPB	7470A		08 Dec 2025 13:01:25
U	Q3770-01MS - 1	4207MS	3.9316	15740	PPB	7470A		08 Dec 2025 13:03:43
U	Q3770-01MSD - 1	4207MSD	3.9891	15973	PPB	7470A		08 Dec 2025 13:05:58
U	CCV75 - 1	CCV75	4.5452	18230	PPB	7470A		08 Dec 2025 13:08:14
U	CCB75 - 1	CCB75	0.0378	-61	PPB	7470A		08 Dec 2025 13:10:29
U	Q3775-01 - 1	NGKO	0.0512	-7	PPB	7470A		08 Dec 2025 13:12:57
U	PB170852BL - 1	PBW	0.0581	21	PPB	7470A		08 Dec 2025 13:15:12
U	PB170852BS - 1	LCSW	3.5075	14019	PPB	7470A		08 Dec 2025 13:17:28
U	Q3776-01 - 1	IDW-RAD-20251203	0.0455	-30	PPB	7470A		08 Dec 2025 13:19:43
U	Q3776-01DUP - 1	IDW-RAD-20251203DUP	0.0689	65	PPB	7470A		08 Dec 2025 13:21:59
U	Q3776-01MS - 1	IDW-RAD-20251203MS	3.7012	14805	PPB	7470A		08 Dec 2025 13:24:16
U	Q3776-01MSD - 1	IDW-RAD-20251203MSD	3.5342	14127	PPB	7470A		08 Dec 2025 13:26:32
U	Q3791-01 - 1	T2	0.0285	-99	PPB	7470A		08 Dec 2025 13:28:49
U	Q3796-01 - 1	Outfall 001	0.0588	24	PPB	7470A		08 Dec 2025 13:31:06
U	Q3796-02 - 1	Outfall 002	0.0637	44	PPB	7470A		08 Dec 2025 13:33:21
U	CCV76 - 1	CCV76	4.9595	19911	PPB	7470A		08 Dec 2025 13:35:37
U	CCB76 - 1	CCB76	0.0341	-76	PPB	7470A		08 Dec 2025 13:37:53
U	Q3796-03 - 1	Outfall 003	0.0366	-66	PPB	7470A		08 Dec 2025 13:40:09
U	Q3792-01 - 1		0.0544	6	PPB	7470A		08 Dec 2025 13:42:25
U	PB170811TB - 1		0.0051	-194	PPB	7470A		08 Dec 2025 13:44:40
U	Q3776-01LX5 - 1		0.0378	-61	PPB	7470A		08 Dec 2025 13:46:56
U	Q3776-01A - 1		3.6857	14742	PPB	7470A		08 Dec 2025 13:49:12
U	Q3770-01LX5 - 1		0.0359	-69	PPB	7470A		08 Dec 2025 13:51:28
U	Q3770-01A - 1		3.7291	14918	PPB	7470A		08 Dec 2025 13:53:45
U	CCV77 - 1	CCV77	4.9058	19693	PPB	7470A		08 Dec 2025 13:56:02
U	CCB77 - 1	CCB77	0.0233	-120	PPB	7470A		08 Dec 2025 13:58:19