

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB138144

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP116007
Turbidity Calibration std, 0NTU	WP116006
Turbidity Calibration std, 20NTU	WP116002
Turbidity Calibration std, 40NTU	WP116001
Turbidity Calibration std, 1NTU	WP116005
Turbidity Calibration std, 5NTU	WP116004
Turbidity Calibration - CCV std, 10 NTU	WP116003
10 NTU Standard 500 ml	W3116

Intercept: 0.2635

Slope: 0.9864

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.165	-0.10		12/08/2025	10:45
2	CAL2	1	1	1.154	0.90	-9.7	12/08/2025	10:48
3	CAL3	5	1	5.460	5.27	5.4	12/08/2025	10:51
4	CAL4	10	1	9.670	9.54	-4.6	12/08/2025	10:54
5	CAL5	20	1	20.600	20.62	3.1	12/08/2025	10:57
6	CAL6	40	1	39.500	39.78	-0.6	12/08/2025	11:00



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB138144

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.276	12/08/2025	11:03
2	ICB		1	0.156	-0.109	12/08/2025	11:06
3	CCV1	10	1	9.840	9.709	12/08/2025	11:09
4	CCB1		1	0.162	-0.103	12/08/2025	11:12
5	LB138144BL		1	0.159	-0.106	12/08/2025	11:15
6	Q3530-12		1	0.728	0.471	12/08/2025	11:18
7	CCV2	10	1	9.780	9.648	12/08/2025	11:21
8	CCB2		1	0.172	-0.093	12/08/2025	11:24