

Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB138146

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/08/2025	11:11
2	CAL2	1	Water	NA	NA	20.2	7.00	12/08/2025	11:12
3	CAL3	1	Water	NA	NA	20.3	10.02	12/08/2025	11:13
4	ICV	1	Water	NA	NA	20.3	7.01	12/08/2025	11:15
5	CCV1	1	Water	NA	NA	20.2	2.02	12/08/2025	11:17
6	Q3776-01	1	Water	NA	NA	20.9	6.72	12/08/2025	11:25
7	Q3776-01DUP	1	Water	NA	NA	21.1	6.74	12/08/2025	11:27
8	CCV2	1	Water	NA	NA	20.2	12.02	12/08/2025	11:30

WORKLIST(Hardcopy Internal Chain)

JB 138146

WorkList Name : corrosivity q3776 w WorkList ID : 193526 Department : Wet-Chemistry Date : 12-08-2025 10:52:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3776-01	D	IDW-RAD-20251203	Water	Corrosivity	Cool 4 deg C	TETR06	A11	12/03/2025 9040C

Date/Time 12/08/25 11:10
Raw Sample Received by: JH wec
Raw Sample Relinquished by: JH wec

Date/Time 12/08/25 13:10
Raw Sample Received by: JH wec
Raw Sample Relinquished by: JH wec