

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138147

Slope : 98.6

BalanceID: MSC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/08/2025	11:11
2	CAL2	1	Water	NA	NA	20.2	7.00	12/08/2025	11:12
3	CAL3	1	Water	NA	NA	20.3	10.02	12/08/2025	11:13
4	ICV	1	Water	NA	NA	20.3	7.01	12/08/2025	11:15
5	CCV1	1	Water	NA	NA	20.1	2.01	12/08/2025	11:36
6	Q3804-01	1	Solid	20.02	20	20.7	7.92	12/08/2025	11:44
7	Q3804-01DUP	1	Solid	20.03	20	20.9	7.94	12/08/2025	11:46
8	Q3804-03	1	Solid	20.02	20	21.1	6.85	12/08/2025	11:55
9	Q3804-04	1	Solid	20.03	20	21.3	6.70	12/08/2025	11:59
10	Q3805-02	1	Solid	20.02	20	20.4	7.29	12/08/2025	12:00
11	CCV2	1	Water	NA	NA	20.3	12.02	12/08/2025	12:02

WORKLIST(Hardcopy Internal Chain)

138147

WorkList Name : ph s q3804 WorkList ID : 193528 Department : Wet-Chemistry Date : 12-08-2025 10:55:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3804-01	120325-A	Solid	pH	Cool 4 deg C	PSEG03	A23	12/05/2025	9045D
Q3804-03	120325-B	Solid	pH	Cool 4 deg C	PSEG03	A23	12/05/2025	9045D
Q3804-04	120325-C	Solid	pH	Cool 4 deg C	PSEG03	A23	12/05/2025	9045D
Q3805-02	91725	Solid	pH	Cool 4 deg C	PSEG03	A31	12/05/2025	9045D

Date/Time 12/08/25 11:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12/08/25 16:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]