

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138149

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/08/2025	11:11
2	CAL2	1	Water	NA	NA	20.2	7.00	12/08/2025	11:12
3	CAL3	1	Water	NA	NA	20.3	10.02	12/08/2025	11:13
4	ICV	1	Water	NA	NA	20.3	7.01	12/08/2025	11:15
5	CCV1	1	Water	NA	NA	20.2	2.01	12/08/2025	12:33
6	Q3791-01	1	Water	NA	NA	20.4	7.26	12/08/2025	12:37
7	Q3791-01DUP	1	Water	NA	NA	20.6	7.28	12/08/2025	12:39
8	Q3792-01	1	Water	NA	NA	20.1	6.55	12/08/2025	12:44
9	Q3792-01DUP	1	Water	NA	NA	20.3	6.57	12/08/2025	12:47
10	Q3793-01	1	Water	NA	NA	20.1	5.83	12/08/2025	12:55
11	Q3793-01DUP	1	Water	NA	NA	20.3	5.85	12/08/2025	12:59
12	Q3805-01	1	Water	NA	NA	20.6	7.82	12/08/2025	13:10
13	Q3805-01DUP	1	Water	NA	NA	20.8	7.84	12/08/2025	13:12
14	CCV2	1	Water	NA	NA	20.2	12.02	12/08/2025	13:15