

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138163

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE (ICV)	W3248

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	12/09/2025	14:00
2	Q3776-01		1	14	100.00	>212.0	>212.0	12/09/2025	14:30
3	Q3791-01		1	12	100.00	>212.0	>212.0	12/09/2025	15:00
4	Q3792-01		1	14	100.00	>212.0	>212.0	12/09/2025	15:30
5	Q3805-01		1	15	100.00	>212.0	>212.0	12/09/2025	16:00
6	Q3805-01DUP		1	15	100.00	>212.0	>212.0	12/09/2025	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6138163

WorkList Name : fp-12-8

WorkList ID : 193509

Department : Wet-Chemistry

Date : 12-08-2025 08:11:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3776-01	IDW-RAD-20251203	Water	Flash Point	Cool 4 deg C	TETR06	A11	12/03/2025	1010B
Q3791-01	T2	Water	Flash Point	Cool 4 deg C	PSEG03	D61	12/05/2025	1010B
Q3792-01	FRAC TANK#251193	Water	Flash Point	Cool 4 deg C	PSEG03	D32	12/05/2025	1010B
Q3805-01	92925	Water	Flash Point	Cool 4 deg C	PSEG03	A31	12/05/2025	1010B

Date/Time 12/19/2025 13:45
 Raw Sample Received by: R. H. W. G.
 Raw Sample Relinquished by: J. H. W. G.

Date/Time 12/19/2025 17:10
 Raw Sample Received by: J. H. W. G.
 Raw Sample Relinquished by: J. H. W. G.