

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB138178

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3264
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/10/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/10/2025	08:01
3	CAL3	1	Water	NA	NA	20.3	10.02	12/10/2025	08:03
4	ICV	1	Water	NA	NA	20.3	7.01	12/10/2025	08:05
5	CCV1	1	Water	NA	NA	20.3	2.01	12/10/2025	08:10
6	Q3811-01	1	Water	NA	NA	20.2	6.88	12/10/2025	08:18
7	Q3811-01DUP	1	Water	NA	NA	20.3	6.90	12/10/2025	08:20
8	Q3811-02	1	Water	NA	NA	20.4	7.13	12/10/2025	08:30
9	Q3811-02DUP	1	Water	NA	NA	20.5	7.14	12/10/2025	08:33
10	Q3811-03	1	Water	NA	NA	20.3	7.38	12/10/2025	08:37
11	Q3811-03DUP	1	Water	NA	NA	20.4	7.40	12/10/2025	08:39
12	Q3822-01	1	Water	NA	NA	20.2	5.67	12/10/2025	08:45
13	Q3822-01DUP	1	Water	NA	NA	20.3	5.69	12/10/2025	08:47
14	Q3824-01	1	Water	NA	NA	20.4	5.98	12/10/2025	08:55
15	Q3824-01DUP	1	Water	NA	NA	20.6	6.00	12/10/2025	08:56
16	CCV2	1	Water	NA	NA	20.3	12.02	12/10/2025	09:00

WORKLIST(Hardcopy Internal Chain)

WP 138178

WorkList Name : ph w q3824 WorkList ID : 193559 Department : Wet-Chemistry Date : 12-10-2025 07:41:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3811-01	C Outfall 001	Water	pH	Cool 4 deg C	DALT03	D31	11/20/2025	9040C
Q3811-02	C Outfall 002	Water	pH	Cool 4 deg C	DALT03	D31	11/20/2025	9040C
Q3811-03	C Outfall 003	Water	pH	Cool 4 deg C	DALT03	D31	11/20/2025	9040C
Q3822-01	A AUD-25-0205	Water	pH	Cool 4 deg C	PSEG03	D31	12/09/2025	9040C
Q3824-01	Y TRE-25-0135	Water	pH	Cool 4 deg C	PSEG03	D31	12/09/2025	9040C

07:50

Date/Time 12/10/25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 12/10/25 13:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]