

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138182

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE (ICV)	W3248

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	83.0	12/10/2025	08:40
2	Q3822-01		1	13	100.00	>212.0	>212.0	12/10/2025	09:10
3	Q3822-01DUP		1	13	100.00	>212.0	>212.0	12/10/2025	09:40
4	Q3824-01		1	15	100.00	>212.0	>212.0	12/10/2025	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66138182

WorkList Name : FP-12-10

WorkList ID : 193572

Department : Wet-Chemistry

Date : 12-10-2025 08:20:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3822-01	AUD-25-0205	Water	Flash Point	Cool 4 deg C	PSEG03	D31	12/09/2025	1010B
Q3824-01	TRE-25-0135	Water	Flash Point	Cool 4 deg C	PSEG03	D31	12/09/2025	1010B

Date/Time 12/10/2025 08:30
 Raw Sample Received by: RH Wey
 Raw Sample Relinquished by: RH Wey

Date/Time 12/10/2025 11:00
 Raw Sample Received by: RH Wey
 Raw Sample Relinquished by: RH Wey