

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB138199

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	12/11/2025	13:45
2	Q3823-01		1	12	100.00	>212.0	>212.0	12/11/2025	14:15
3	Q3823-01DUP		1	12	100.00	>212.0	>212.0	12/11/2025	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6138199

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-3823

WorkList ID : 193607

Department : Wet-Chemistry

Date : 12-11-2025 08:22:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3823-01	BUR-1321	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	12/09/2025	1010B

Date/Time 12/11/2025 13:35
Raw Sample Received by: RM (w/)
Raw Sample Relinquished by: JG (w/)

Date/Time 12/11/2025 15:25
Raw Sample Received by: JG (w/)
Raw Sample Relinquished by: RM (w/)