

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB138205

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP116074
Chlorine Calibration std, 0.1ppm	WP116069
Chlorine Calibration std, 0.2ppm	WP116070
Chlorine Calibration std, 0.8ppm	WP116072
Chlorine Calibration std, 0.0ppm	WP116068
Chlorine Calibration std, 1.6ppm	WP116073
Residual Chlorine Calibration and CCV std, 0	WP116071
Total Chlorine Powder Pillows	W3147

Intercept: 0.0094

Slope: 0.9785

Regression: 0.999697

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		12/11/2025	15:50
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	13	12/11/2025	15:53
3	CAL3	0.2	1	0.000	0.210	0.210	0.21	2.5	12/11/2025	15:56
4	CAL4	0.4	1	0.000	0.380	0.380	0.38	-5.3	12/11/2025	15:59
5	CAL5	0.8	1	0.000	0.810	0.810	0.82	2.2	12/11/2025	16:02
6	CAL6	1.6	1	0.000	1.570	1.570	1.60	-0.3	12/11/2025	16:05

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB138205

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4090	12/11/2025	16:08
2	ICB				1	0.0000	0.0100	0.0100	0.0010	12/11/2025	16:11
3	CCV1			0.4	1	0.0000	0.3900	0.3900	0.3890	12/11/2025	16:14
4	CCB1				1	0.0000	0.0100	0.0100	0.0010	12/11/2025	16:17
5	LB138205BL	50	50		1	0.0000	0.0000	0.0000	-0.0100	12/11/2025	16:20
6	LB138205BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4200	12/11/2025	16:23
7	Q3841-01	50	50		1	0.0000	0.1600	0.1600	0.1540	12/11/2025	16:26
8	Q3841-01DUP	50	50		1	0.0000	0.1500	0.1500	0.1440	12/11/2025	16:29
9	Q3841-01MS	50	50	0.4	1	0.0000	0.5300	0.5300	0.5320	12/11/2025	16:32
10	Q3841-01MSD	50	50	0.4	1	0.0000	0.5400	0.5400	0.5420	12/11/2025	16:35
11	Q3841-02	50	50		1	0.0000	0.2000	0.2000	0.1950	12/11/2025	16:38
12	Q3841-03	50	50		1	0.0000	0.0200	0.0200	0.0110	12/11/2025	16:41
13	Q3841-04	50	50		1	0.0000	0.0500	0.0500	0.0410	12/11/2025	16:44
14	Q3841-05	50	50		1	0.0000	0.0300	0.0300	0.0210	12/11/2025	16:47
15	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3990	12/11/2025	16:50
16	CCB2				1	0.0000	0.0100	0.0100	0.0010	12/11/2025	16:53
17	Q3841-06	50	50		1	0.0000	0.0300	0.0300	0.0210	12/11/2025	16:57
18	Q3841-07	50	50		1	0.0000	0.0600	0.0600	0.0520	12/11/2025	17:00
19	CCV3			0.4	1	0.0000	0.4000	0.4000	0.3990	12/11/2025	17:03
20	CCB3				1	0.0000	0.0100	0.0100	0.0010	12/11/2025	17:06

WORKLIST(Hardcopy Internal Chain)

LB138205

WorkList Name : RESCHLORINE-121125 WorkList ID : 193608 Department : Wet-Chemistry Date : 12-11-2025 15:59:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3841-01	B2-W14-RM3088-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G
Q3841-02	B2-W14-RM3075-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G
Q3841-03	B2-W14-LHDF	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G
Q3841-04	B2-W14-RM3055-RH-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G
Q3841-05	B2-W14-EW	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G
Q3841-06	B2-W14-RM3056-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G
Q3841-07	B2-W14-RM3073-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	D41	12/11/2025	SM4500 Cl G

Date/Time 12/11/25 15:40
 Raw Sample Received by: 12(wc)
 Raw Sample Relinquished by: 28(wc)

Date/Time 12/11/25 17:20
 Raw Sample Received by: 28(wc)
 Raw Sample Relinquished by: 12(wc)