

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB138214

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3264
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/15/2025	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	12/15/2025	08:06
3	CAL3	1	Water	NA	NA	20.2	10.02	12/15/2025	08:08
4	ICV	1	Water	NA	NA	20.2	6.99	12/15/2025	08:10
5	CCV1	1	Water	NA	NA	20.2	2.01	12/15/2025	08:11
6	Q3856-02	1	Water	NA	NA	20.1	7.70	12/15/2025	08:20
7	Q3856-02DUP	1	Water	NA	NA	20.3	7.72	12/15/2025	08:22
8	Q3857-01	1	Water	NA	NA	20.4	7.90	12/15/2025	08:30
9	Q3857-01DUP	1	Water	NA	NA	20.6	7.92	12/15/2025	08:33
10	Q3857-02	1	Water	NA	NA	20.3	7.07	12/15/2025	08:40
11	Q3857-02DUP	1	Water	NA	NA	20.4	7.09	12/15/2025	08:44
12	Q3857-03	1	Water	NA	NA	20.9	4.88	12/15/2025	08:55
13	Q3857-03DUP	1	Water	NA	NA	20.8	4.90	12/15/2025	09:00
14	Q3857-04	1	Water	NA	NA	20.7	6.45	12/15/2025	09:12
15	Q3857-04DUP	1	Water	NA	NA	20.8	6.47	12/15/2025	09:14
16	CCV2	1	Water	NA	NA	20.2	12.02	12/15/2025	09:15

WORKLIST(Hardcopy Internal Chain)

138214

WorkList Name : ph w q3857 WorkList ID : 193638 Department : Wet-Chemistry Date : 12-15-2025 07:49:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3856-02	B 327	Water	pH	Cool 4 deg C	PSEG03	D31	12/12/2025	9040C
Q3857-01	A 1117	Water	pH	Cool 4 deg C	PSEG03	H21	12/12/2025	9040C
Q3857-02	A 34934	Water	pH	Cool 4 deg C	PSEG03	H21	12/12/2025	9040C
Q3857-03	A 1106	Water	pH	Cool 4 deg C	PSEG03	H21	12/12/2025	9040C
Q3857-04	A 34935	Water	pH	Cool 4 deg C	PSEG03	H21	12/12/2025	9040C

Date/Time 12-15-25 08:00
 Raw Sample Received by: JG WDC
 Raw Sample Relinquished by: R3 L E-L-1259

Date/Time 12-15-25 13:38
 Raw Sample Received by: R3 L E-L-1259
 Raw Sample Relinquished by: JG WDC