

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB138228

Slope : 98.6

BalanceID: MSC-3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3264
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/16/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/16/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	12/16/2025	08:14
4	ICV	1	Water	NA	NA	20.2	7.01	12/16/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	12/16/2025	08:20
6	Q3871-01	1	Solid	20.02	20	20.1	5.30	12/16/2025	08:30
7	Q3871-03	1	Solid	20.03	20	20.1	5.51	12/16/2025	08:33
8	Q3871-05	1	Solid	20.04	20	20.8	5.97	12/16/2025	08:40
9	Q3872-01	1	Solid	20.02	20	20.1	8.39	12/16/2025	08:55
10	Q3872-03	1	Solid	20.03	20	20.4	7.17	12/16/2025	09:00
11	Q3873-02	1	Solid	20.02	20	20.1	6.66	12/16/2025	09:10
12	Q3873-04	1	Solid	20.03	20	20.1	6.70	12/16/2025	09:15
13	Q3880-01	1	Solid	20.02	20	20.8	6.99	12/16/2025	09:25
14	Q3881-03	1	Solid	20.04	20	20.6	11.72	12/16/2025	09:37
15	Q3881-03DUP	1	Solid	20.03	20	20.8	11.74	12/16/2025	09:40
16	CCV2	1	Water	NA	NA	20.2	12.02	12/16/2025	09:41

WORKLIST(Hardcopy Internal Chain)

138228

WorkList Name : corrsivity q3871 WorkList ID : 193667 Department : Wet-Chemistry Date : 12-16-2025 07:38:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3871-01	LAW-25-0194	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	12/12/2025	9045D
Q3871-03	LAW-25-0195	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	12/12/2025	9045D
Q3871-05	LAW-25-0192-0193-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	12/12/2025	9045D
Q3872-01	ARS20-0006	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	12/12/2025	9045D
Q3872-03	MOO-25-0367	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	12/12/2025	9045D
Q3873-02	COMP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	12/12/2025	9045D
Q3873-04	COMP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	12/12/2025	9045D
Q3880-01	BG-SET-2	Solid	Corrosivity	Cool 4 deg C	PSEG04	D31	12/15/2026	9045D
Q3881-03	DM-X-76	Solid	Corrosivity	Cool 4 deg C	JPCL01	D31	12/15/2025	9045D

Date/Time 12-16-25 08:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12-16-25 13:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]