

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138229

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3264
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/16/2025	09:25
2	CAL2	1	Water	NA	NA	20.2	7.01	12/16/2025	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	12/16/2025	09:30
4	ICV	1	Water	NA	NA	20.2	7.01	12/16/2025	09:33
5	CCV1	1	Water	NA	NA	20.2	2.02	12/16/2025	09:35
6	Q3880-02	1	Water	NA	NA	20.4	7.07	12/16/2025	09:44
7	Q3880-02DUP	1	Water	NA	NA	20.5	7.08	12/16/2025	09:47
8	CCV2	1	Water	NA	NA	20.3	12.02	12/16/2025	09:50

WORKLIST(Hardcopy Internal Chain)

136229

WorkList Name : ph w q3880 WorkList ID : 193682 Department : Wet-Chemistry Date : 12-16-2025 09:08:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3880-02	FB	Water	pH	Cool 4 deg C	PSEG04	D31	12/15/2026	9040C

Date/Time 12/16/25 09:15
 Raw Sample Received by: SA WCI
 Raw Sample Relinquished by: ACR

Date/Time 12/16/25 11:40
 Raw Sample Received by: CJM
 Raw Sample Relinquished by: SA WCI