

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB138243

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3264
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/17/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/17/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	12/17/2025	08:15
4	ICV	1	Water	NA	NA	20.2	7.00	12/17/2025	08:19
5	CCV1	1	Water	NA	NA	20.2	2.01	12/17/2025	08:20
6	Q3894-02	1	Water	NA	NA	20.4	5.08	12/17/2025	08:30
7	Q3894-02DUP	1	Water	NA	NA	20.6	5.09	12/17/2025	08:32
8	CCV2	1	Water	NA	NA	20.3	12.02	12/17/2025	08:35

WORKLIST(Hardcopy Internal Chain)

VF 138243

WorkList Name : ph q3894 WorkList ID : 193705 Department : Wet-Chemistry Date : 12-17-2025 07:44:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3894-02	C 120425	Water	pH	Cool 4 deg C	PSEG03	D31	12/16/2025	9040C

Date/Time 12-17-25 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 12-17-25 12:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]