

CB131334


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

41890



A= 0.0000e+000

B= 2.4105e-004

C= 2.0434e-002

Rho= 0.9997613

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.069	0.069	201	0.000	201	0				2
0.2	0.200	0.204	0.004	762	0.0 %	762					-1
2.5	2.500	2.469	-0.031	10158	0.0 %	10158					0
5.0	5.000	4.977	-0.023	20562	0.0 %	20562					-2
7.5	7.500	7.363	-0.137	30460	0.0 %	30460					1
10.0	10.000	10.118	0.118	41890	0.0 %	41890					

LB131334 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jun 2024 14:35:40

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	201	-	0.0000	7470A	PPB	20 Jun 2024 14:39:53	S	Std
0.2 - 1	542	762	-	0.2000	7470A	PPB	20 Jun 2024 14:42:10	S	Std
2.5 - 1	525	10158	-	2.5000	7470A	PPB	20 Jun 2024 14:44:28	S	Std
5.0 - 1	550	20562	-	5.0000	7470A	PPB	20 Jun 2024 14:46:44	S	Std
7.5 - 1	575	30460	-	7.5000	7470A	PPB	20 Jun 2024 14:49:02	S	Std
10.0 - 1	518	41890	-	10.0000	7470A	PPB	20 Jun 2024 14:51:21	S	Std
ICV97 - 1	ICV97	15511	3.7594	-	7470A	PPB	20 Jun 2024 14:54:11	U	SMPL
ICB97 - 1	ICB97	-31	0.0130	-	7470A	PPB	20 Jun 2024 14:56:28	U	SMPL
CCV64 - 1	CCV64	20309	4.9160	-	7470A	PPB	20 Jun 2024 14:58:45	U	SMPL
CCB64 - 1	CCB64	-38	0.0113	-	7470A	PPB	20 Jun 2024 15:01:00	U	SMPL
CRA - 1	CRA	620	0.1699	-	7470A	PPB	20 Jun 2024 15:03:17	U	SMPL
HighStd - 1	HighStd	41769	10.0889	-	7470A	PPB	20 Jun 2024 15:05:32	U	SMPL
ChkStd - 1	ChkStd	28533	6.8984	-	7470A	PPB	20 Jun 2024 15:07:47	U	SMPL
PB161613BL - 1	PBW	-155	-0.0169	-	7470A	PPB	20 Jun 2024 15:10:06	U	SMPL
PB161613BS - 1	LCSW	13777	3.3414	-	7470A	PPB	20 Jun 2024 15:12:24	U	SMPL
P2942-01 - 1	GRAY-WATER	20340	4.9234	-	7470A	PPB	20 Jun 2024 15:14:40	U	SMPL
P2942-01DUP - 1	GRAY-WATERDUP	17498	4.2384	-	7470A	PPB	20 Jun 2024 15:16:59	U	SMPL
P2942-01MS - 1	GRAY-WATERMS	33602	8.1203	-	7470A	PPB	20 Jun 2024 15:19:19	U	SMPL
P2942-01MSD - 1	GRAY-WATERMSD	34031	8.2237	-	7470A	PPB	20 Jun 2024 15:21:38	U	SMPL
CCV65 - 1	CCV65	19162	4.6395	-	7470A	PPB	20 Jun 2024 15:23:55	U	SMPL
CCB65 - 1	CCB65	-398	-0.0755	-	7470A	PPB	20 Jun 2024 15:26:15	U	SMPL
P2942-01LX5 - 1		3821	0.9415	-	7470A	PPB	20 Jun 2024 15:28:33	U	SMPL
P2942-01A - 1		31794	7.6845	-	7470A	PPB	20 Jun 2024 15:30:48	U	SMPL
CCV67 - 1	CCV67	19511	4.7236	-	7470A	PPB	20 Jun 2024 15:34:45	U	SMPL
CCB67 - 1	CCB67	-216	-0.0316	-	7470A	PPB	20 Jun 2024 15:37:03	U	SMPL