

LB99074


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42562



A= 0.0000e+000

B= 2.3509e-004

C= -7.2261e-002

Rho= 0.9998455

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	266	0.000	266				
0.2	0.200	0.176	-0.024	1056	0.0 %	1056				
2.5	2.500	2.454	-0.046	10748	0.0 %	10748				
5.0	5.000	5.134	0.134	22146	0.0 %	22146				
7.5	7.500	7.512	0.012	32259	0.0 %	32259				
10.0	10.000	9.934	-0.066	42562	0.0 %	42562				

LB99074 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Nov 2018 10:34:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	266	-	0.0000	7470A	PPB	07 Nov 2018 10:37:29	S	Std
0.2 - 1	50.2	1056	-	0.2000	7470A	PPB	07 Nov 2018 10:39:38	S	Std
2.5 - 1	52.5	10748	-	2.5000	7470A	PPB	07 Nov 2018 10:50:10	S	Std
5.0 - 1	55.0	22146	-	5.0000	7470A	PPB	07 Nov 2018 10:52:18	S	Std
7.5 - 1	57.5	32259	-	7.5000	7470A	PPB	07 Nov 2018 10:54:26	S	Std
10.0 - 1	510.0	42562	-	10.0000	7470A	PPB	07 Nov 2018 11:02:15	S	Std
ICV62 - 1	ICV62	16409	3.7853	-	7470A	PPB	07 Nov 2018 11:09:28	U	SMPL
ICB62 - 1	ICB62	210	-0.0229	-	7470A	PPB	07 Nov 2018 11:11:34	U	SMPL
CCV22 - 1	CCV22	22209	5.1489	-	7470A	PPB	07 Nov 2018 11:13:41	U	SMPL
CCB22 - 1	CCB22	193	-0.0269	-	7470A	PPB	07 Nov 2018 11:15:47	U	SMPL
CRA - 1	CRA	1002	0.1633	-	7470A	PPB	07 Nov 2018 11:17:55	U	SMPL
HighStd - 1	HighStd	41726	9.7371	-	7470A	PPB	07 Nov 2018 11:20:01	U	SMPL
ChkStd - 1	ChkStd	30313	7.0540	-	7470A	PPB	07 Nov 2018 11:22:08	U	SMPL
PB114569BL - 1	PBW	148	-0.0375	-	7470A	PPB	07 Nov 2018 11:26:47	U	SMPL
PB114569BS - 1	LCSW	15030	3.4611	-	7470A	PPB	07 Nov 2018 11:31:21	U	SMPL
J5765-02 - 1	NB-08-110518	4	-0.0713	-	7470A	PPB	07 Nov 2018 11:33:28	U	SMPL
J5765-02DUP - 1	NB-08-110518DUP	69	-0.0560	-	7470A	PPB	07 Nov 2018 11:35:36	U	SMPL
J5765-02MS - 1	NB-08-110518MS	15613	3.5982	-	7470A	PPB	07 Nov 2018 11:40:13	U	SMPL
J5765-02MSD - 1	NB-08-110518MSD	15573	3.5888	-	7470A	PPB	07 Nov 2018 11:42:21	U	SMPL
J5772-02 - 1	PL-05-110518	-24	-0.0779	-	7470A	PPB	07 Nov 2018 11:44:28	U	SMPL
CCV23 - 1	CCV23	22160	5.1373	-	7470A	PPB	07 Nov 2018 11:48:57	U	SMPL
CCB23 - 1	CCB23	169	-0.0325	-	7470A	PPB	07 Nov 2018 11:51:03	U	SMPL
J5819-09 - 1	BES48_WC	109	-0.0466	-	7470A	PPB	07 Nov 2018 11:53:10	U	SMPL
J5826-04 - 1	SB-1	53	-0.0598	-	7470A	PPB	07 Nov 2018 11:55:17	U	SMPL
J5826-08 - 1	SB-2	-15	-0.0758	-	7470A	PPB	07 Nov 2018 11:57:24	U	SMPL
J5826-12 - 1	SB-S	325	0.0041	-	7470A	PPB	07 Nov 2018 11:59:31	U	SMPL
J5827-10 - 1	SP-A	281	-0.0062	-	7470A	PPB	07 Nov 2018 12:01:38	U	SMPL
J5827-20 - 1	SP-B	124	-0.0431	-	7470A	PPB	07 Nov 2018 12:03:45	U	SMPL
J5827-30 - 1	SP-C	62	-0.0577	-	7470A	PPB	07 Nov 2018 12:05:53	U	SMPL
J5827-40 - 1	SP-D	145	-0.0382	-	7470A	PPB	07 Nov 2018 12:08:02	U	SMPL
J5827-50 - 1	SP-E	1	-0.0720	-	7470A	PPB	07 Nov 2018 12:10:11	U	SMPL
PB114570BL - 1	PBW	286	-0.0050	-	7470A	PPB	07 Nov 2018 12:12:18	U	SMPL
CCV24 - 1	CCV24	21455	4.9716	-	7470A	PPB	07 Nov 2018 12:16:56	U	SMPL
CCB24 - 1	CCB24	308	0.0001	-	7470A	PPB	07 Nov 2018 12:19:02	U	SMPL
PB114570BS - 1	LCSW	16653	3.8427	-	7470A	PPB	07 Nov 2018 12:23:34	U	SMPL
J5820-01 - 1	CPS043085	-100	-0.0958	-	7470A	PPB	07 Nov 2018 12:25:41	U	SMPL
J5820-03 - 1	CPS093081	-31	-0.0795	-	7470A	PPB	07 Nov 2018 12:27:48	U	SMPL
J5820-05 - 1	CPS043081	-51	-0.0843	-	7470A	PPB	07 Nov 2018 12:29:56	U	SMPL
J5820-07 - 1	C19013081	-43	-0.0824	-	7470A	PPB	07 Nov 2018 12:32:03	U	SMPL
J5820-07DUP - 1	C19013081DUP	-44	-0.0826	-	7470A	PPB	07 Nov 2018 12:34:11	U	SMPL
J5820-09 - 1	J5820-07MS	16566	3.8222	-	7470A	PPB	07 Nov 2018 12:39:06	U	SMPL
J5820-11 - 1	J5820-07MSD	16486	3.8034	-	7470A	PPB	07 Nov 2018 12:41:14	U	SMPL
J5820-13 - 1	CPS083081	-64	-0.0873	-	7470A	PPB	07 Nov 2018 12:43:22	U	SMPL
J5820-15 - 1	CPS103081	7	-0.0706	-	7470A	PPB	07 Nov 2018 12:45:30	U	SMPL
CCV25 - 1	CCV25	21107	4.8898	-	7470A	PPB	07 Nov 2018 12:47:37	U	SMPL
CCB25 - 1	CCB25	102	-0.0483	-	7470A	PPB	07 Nov 2018 12:49:44	U	SMPL
J5820-17 - 1	CPS123081	45	-0.0617	-	7470A	PPB	07 Nov 2018 12:51:51	U	SMPL
J5822-01 - 1	FB-01	-30	-0.0793	-	7470A	PPB	07 Nov 2018 12:53:59	U	SMPL
J5822-03 - 1	MW-7	9	-0.0701	-	7470A	PPB	07 Nov 2018 12:58:13	U	SMPL
J5822-04 - 1	MW-8	-24	-0.0779	-	7470A	PPB	07 Nov 2018 13:00:20	U	SMPL
J5822-05 - 1	MW-9	176	-0.0309	-	7470A	PPB	07 Nov 2018 13:02:28	U	SMPL
J5822-06 - 1	MW-11	35	-0.0640	-	7470A	PPB	07 Nov 2018 13:04:36	U	SMPL
PB114571BL - 1	PBW	210	-0.0229	-	7470A	PPB	07 Nov 2018 13:08:53	U	SMPL
PB114571BS - 1	LCSW	16536	3.8152	-	7470A	PPB	07 Nov 2018 13:11:00	U	SMPL
CCV26 - 1	CCV26	21685	5.0257	-	7470A	PPB	07 Nov 2018 13:13:07	U	SMPL
CCB26 - 1	CCB26	80	-0.0535	-	7470A	PPB	07 Nov 2018 13:15:14	U	SMPL
J5820-02 - 1	CPS043085	-39	-0.0814	-	7470A	PPB	07 Nov 2018 13:17:22	U	SMPL
J5820-04 - 1	CPS093081	-3	-0.0730	-	7470A	PPB	07 Nov 2018 13:19:29	U	SMPL
J5820-06 - 1	CPS043081	-27	-0.0786	-	7470A	PPB	07 Nov 2018 13:21:37	U	SMPL
J5820-08 - 1	C19013081	-30	-0.0793	-	7470A	PPB	07 Nov 2018 13:23:44	U	SMPL
J5820-08DUP - 1	C19013081DUP	-22	-0.0774	-	7470A	PPB	07 Nov 2018 13:25:51	U	SMPL
J5820-10 - 1	J5820-08MS	16401	3.7835	-	7470A	PPB	07 Nov 2018 13:27:59	U	SMPL
J5820-12 - 1	J5820-08MSD	16689	3.8512	-	7470A	PPB	07 Nov 2018 13:30:07	U	SMPL
J5820-14 - 1	CPS083081	-86	-0.0925	-	7470A	PPB	07 Nov 2018 13:32:16	U	SMPL
J5820-16 - 1	CPS103081	-19	-0.0767	-	7470A	PPB	07 Nov 2018 13:34:24	U	SMPL
J5820-18 - 1	CPS123081	-56	-0.0854	-	7470A	PPB	07 Nov 2018 13:36:32	U	SMPL
CCV27 - 1	CCV27	21282	4.9309	-	7470A	PPB	07 Nov 2018 13:38:39	U	SMPL
CCB27 - 1	CCB27	185	-0.0288	-	7470A	PPB	07 Nov 2018 13:40:46	U	SMPL
PB114569TB - 1	PB114530TB	127	-0.0424	-	7470A	PPB	07 Nov 2018 13:42:53	U	SMPL
J5822-02 - 1	MW-6	1468	0.2729	-	7470A	PPB	07 Nov 2018 13:45:01	U	SMPL
J5822-07 - 1	DUP-1	1183	0.2059	-	7470A	PPB	07 Nov 2018 13:47:08	U	SMPL
J5765-02LX5 - 1		238	-0.0163	-	7470A	PPB	07 Nov 2018 13:49:15	U	SMPL
J5765-02A - 1		15512	3.5745	-	7470A	PPB	07 Nov 2018 13:51:23	U	SMPL
J5820-07LX5 - 1		189	-0.0278	-	7470A	PPB	07 Nov 2018 13:53:31	U	SMPL
J5820-07A - 1		15309	3.5267	-	7470A	PPB	07 Nov 2018 13:55:40	U	SMPL
J5820-08LX5 - 1		191	-0.0274	-	7470A	PPB	07 Nov 2018 13:57:48	U	SMPL
J5820-08A - 1		15414	3.5514	-	7470A	PPB	07 Nov 2018 13:59:58	U	SMPL
CCV28 - 1	CCV28	21051	4.8766	-	7470A	PPB	07 Nov 2018 14:02:06	U	SMPL
CCB28 - 1	CCB28	86	-0.0520	-	7470A	PPB	07 Nov 2018 14:04:13	U	SMPL